

Data Path : Z:\HPCHEM1\BNA E\DATA\BE051917\
 Data File : BE093055.D
 Acq On : 19 May 2017 11:28
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 19 17:51:44 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE051517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 15 13:35:57 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	72	0.00
2	1,4-Dioxane	0.810	0.812	-0.2	85	0.00
3	n-Nitrosodimethylamine	0.724	0.767	-5.9	87	0.00
4 S	2-Fluorophenol	1.226	1.289	-5.1	85	0.00
5 S	Phenol-d6	1.656	1.790	-8.1	89	0.00
6	bis(2-Chloroethyl)ether	1.349	1.397	-3.6	85	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	76	0.00
8 S	Nitrobenzene-d5	0.350	0.370	-5.7	93	0.00
9	Naphthalene	1.054	1.029	2.4	83	0.00
10	Hexachlorobutadiene	0.138	0.135	2.2	84	0.00
11 SURR	2-Methylnaphthalene-d10	0.586	0.594	-1.4	86	0.00
12	2-Methylnaphthalene	0.653	0.654	-0.2	86	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	80	0.00
14 S	2,4,6-Tribromophenol	0.118	0.133	-12.7	104	0.00
15 S	2-Fluorobiphenyl	1.645	1.614	1.9	87	0.00
16	Acenaphthylene	7.015	6.942	1.0	93	0.00
17	Acenaphthene	1.263	1.256	0.6	90	0.00
18	Fluorene	1.545	1.596	-3.3	93	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	79	0.00
20	4-Bromophenyl-phenylether	0.124	0.116	6.5	93	0.00
21	Hexachlorobenzene	0.163	0.162	0.6	99	0.00
22	Pentachlorophenol	0.050	0.043	14.0	102	0.00
23	Phenanthrene	0.867	0.815	6.0	95	0.03
24	Anthracene	0.816	0.807	1.1	107	0.00
25 SURR	Fluoranthene-d10	1.051	1.094	-4.1	102	0.00
26	Fluoranthene	4.922	5.094	-3.5	104	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	96	-0.02
28	Pvrene	4.819	4.841	-0.5	105	0.00
29 S	Terphenyl-d14	0.747	0.713	4.6	100	0.00
30	Benzo(a)anthracene	1.088	1.103	-1.4	109	0.00
31	Chrysene	1.163	1.130	2.8	105	0.00
32	Bis(2-ethylhexyl)phthalate	0.627	0.652	-4.0	142	0.00
33	Indeno(1,2,3-cd)pyrene	4.256	4.084	4.0	104	0.00
34 I	Perylene-d12	1.000	1.000	0.0	96	-0.01
35	Benzo(b)fluoranthene	1.264	1.259	0.4	112	0.00
36	Benzo(k)fluoranthene	1.268	1.188	6.3	102	0.00
37 C	Benzo(a)pyrene	1.157	1.121	3.1	109	-0.01
38	Dibenzo(a,h)anthracene	1.076	1.018	5.4	105	0.00
39	Benzo(a,h,i)perylene	4.686	4.556	2.8	107	0.00

Data Path : Z:\HPCHEM1\BNA E\DATA\BE051917\
Data File : BE093055.D
Acq On : 19 May 2017 11:28
Operator : SJ/MA
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 19 17:51:44 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE051517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 15 13:35:57 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			