

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE052316\
 Data File : BE091842.D
 Acq On : 23 May 2016 11:38
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 24 02:31:06 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE051916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 19 15:36:49 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	-0.02
2	1,4-Dioxane	0.649	0.756	-16.5	113	-0.02
3	n-Nitrosodimethylamine	0.798	0.824	-3.3	116	0.00
4 S	2-Fluorophenol	1.307	1.510	-15.5	126	0.00
5 S	Phenol-d6	1.840	1.791	2.7	111	0.00
6	bis(2-Chloroethyl)ether	1.477	1.489	-0.8	113	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	92	-0.02
8 S	Nitrobenzene-d5	0.331	0.359	-8.5	112	-0.01
9	Nitrobenzene	0.341	0.356	-4.4	111	-0.01
10	Naphthalene	1.106	1.034	6.5	90	0.00
11	Hexachlorobutadiene	0.164	0.160	2.4	94	-0.02
12	2-Methylnaphthalene	0.744	0.694	6.7	91	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	86	-0.01
14 S	2,4,6-Tribromophenol	0.194	0.274	-41.2#	140	-0.01
15 S	2-Fluorobiphenyl	1.414	1.407	0.5	90	-0.01
16	Acenaphthylene	2.230	1.946	12.7	79	-0.01
17	Acenaphthene	1.341	1.217	9.2	83	-0.01
18	Fluorene	1.738	1.692	2.6	89	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	99	-0.01
20	4-Bromophenyl-phenylether	0.187	0.168	10.2	91	0.00
21	Hexachlorobenzene	0.195	0.167	14.4	86	0.00
22	Pentachlorophenol	0.080	0.055	31.3#	76	0.00
23	Phenanthrene	1.190	1.054	11.4	87	-0.01
24	Anthracene	1.088	0.990	9.0	93	-0.01
25	Fluoranthene	1.329	1.341	-0.9	100	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	108	-0.02
27	Benzydine	0.338	0.383	-13.3	160#	-0.02
28	Pyrene	1.097	0.968	11.8	100	-0.02
29 S	Terphenyl-d14	0.708	0.691	2.4	114	-0.02
30	Benzo(a)anthracene	1.206	1.260	-4.5	121	-0.02
31	3,3'-Dichlorobenzidine	0.767	0.773	-0.8	125	-0.02
32	Chrysene	1.101	1.007	8.5	106	-0.02
33	Bis(2-ethylhexyl)phthalate	0.540	0.576	-6.7	146	-0.02
34	Indeno(1,2,3-cd)pyrene	1.236	1.227	0.7	114	-0.03
35 I	Perylene-d12	1.000	1.000	0.0	109	-0.02
36	Benzo(b)fluoranthene	1.251	1.245	0.5	115	-0.02
37	Benzo(k)fluoranthene	1.281	1.249	2.5	112	-0.02
38 C	Benzo(a)pyrene	1.194	1.198	-0.3	116	-0.02
39	Dibenzo(a,h)anthracene	1.166	1.147	1.6	115	-0.05
40	Benzo(g,h,i)perylene	1.203	1.166	3.1	112	-0.05

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE052316\
Data File : BE091842.D
Acq On : 23 May 2016 11:38
Operator : UM/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 24 02:31:06 2016
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE051916.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 19 15:36:49 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				