

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE052316\
 Data File : BE091852.D
 Acq On : 23 May 2016 18:06
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 24 05:23:40 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE051916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 19 15:36:49 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	114	-0.02
2	1,4-Dioxane	0.649	0.607	6.5	100	-0.02
3	n-Nitrosodimethylamine	0.798	0.743	6.9	115	0.00
4 S	2-Fluorophenol	1.307	1.298	0.7	120	0.00
5 S	Phenol-d6	1.840	1.827	0.7	126	0.00
6	bis(2-Chloroethyl)ether	1.477	1.465	0.8	123	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	119	-0.02
8 S	Nitrobenzene-d5	0.331	0.337	-1.8	136	-0.01
9	Nitrobenzene	0.341	0.336	1.5	136	-0.01
10	Naphthalene	1.106	1.042	5.8	117	0.00
11	Hexachlorobutadiene	0.164	0.157	4.3	119	-0.02
12	2-Methylnaphthalene	0.744	0.715	3.9	121	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	114	-0.01
14 S	2,4,6-Tribromophenol	0.194	0.180	7.2	122	-0.01
15 S	2-Fluorobiphenyl	1.414	1.370	3.1	116	-0.01
16	Acenaphthylene	2.230	2.248	-0.8	121	-0.01
17	Acenaphthene	1.341	1.252	6.6	113	-0.01
18	Fluorene	1.738	1.636	5.9	114	-0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	121	-0.01
20	4-Bromophenyl-phenylether	0.187	0.171	8.6	114	-0.02
21	Hexachlorobenzene	0.195	0.172	11.8	109	0.00
22	Pentachlorophenol	0.080	0.048	40.0#	80	0.00
23	Phenanthrene	1.190	1.001	15.9	101	-0.01
24	Anthracene	1.088	0.954	12.3	110	-0.01
25	Fluoranthene	1.329	1.105	16.9	101	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	100	-0.02
27	Benzydine	0.338	0.481	-42.3#	186#	-0.02
28	Pyrene	1.097	1.066	2.8	102	-0.02
29 S	Terphenyl-d14	0.708	0.777	-9.7	118	-0.02
30	Benzo(a)anthracene	1.206	1.320	-9.5	118	-0.02
31	3,3'-Dichlorobenzidine	0.767	0.701	8.6	105	-0.02
32	Chrysene	1.101	1.004	8.8	98	-0.02
33	Bis(2-ethylhexyl)phthalate	0.540	0.785	-45.4#	183#	-0.02
34	Indeno(1,2,3-cd)pyrene	1.236	1.175	4.9	101	-0.05
35 I	Perylene-d12	1.000	1.000	0.0	96	-0.02
36	Benzo(b)fluoranthene	1.251	1.320	-5.5	107	-0.02
37	Benzo(k)fluoranthene	1.281	1.186	7.4	94	-0.02
38 C	Benzo(a)pyrene	1.194	1.213	-1.6	103	-0.02
39	Dibenzo(a,h)anthracene	1.166	1.174	-0.7	103	-0.06
40	Benzo(g,h,i)perylene	1.203	1.197	0.5	101	-0.06

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE052316\
Data File : BE091852.D
Acq On : 23 May 2016 18:06
Operator : UM/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 24 05:23:40 2016
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE051916.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 19 15:36:49 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				