

Data Path : Z:\HPCHEM1\BNA E\DATA\BE052516\
 Data File : BE091866.D
 Acq On : 25 May 2016 19:50
 Operator : UM/SJ
 Sample : H3065-04
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SS043MW010-160511

Manual Integrations
 APPROVED

umangi
 6/22/2016 5:40:08 PM

Quant Time: May 26 01:48:00 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE052516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 25 19:42:30 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.78	152	32944	5.00	ng	0.00
8) Naphthalene-d8	10.55	136	174080	5.00	ng	0.00
15) Acenaphthene-d10	14.40	164	111041	5.00	ng	0.00
24) Phenanthrene-d10	17.14	188	269241	5.00	ng	0.00
31) Chrysene-d12	21.29	240	330187	5.00	ng	0.00
40) Perylene-d12	23.71	264	286588	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.39	112	14896	1.58	ng	0.00
5) Phenol-d6	6.95	99	15763	1.28	ng	0.00
9) Nitrobenzene-d5	8.91	82	30588	2.16	ng	-0.02
16) 2,4,6-Tribromophenol	15.87	330	30663	6.19	ng	-0.02
17) 2-Fluorobiphenyl	13.01	172	91266	2.76	ng	-0.02
34) Terphenyl-d14	19.74	244	181423	3.64	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.31	88	8172	1.98	ng	Qvalue 95
6) bis(2-Chloroethyl)ether	7.20	93	3306	0.32	ng	96
12) Naphthalene	10.61	128	1796	0.05	ng	# 68
20) Acenaphthene	14.44	154	185941	7.10	ng	# 83
22) Fluorene	15.44	166	195232	5.01	ng	99
27) Pentachlorophenol	16.78	266	492	0.14	ng	93
28) Phenanthrene	17.17	178	83682	1.61	ng	100
29) Anthracene	17.26	178	49232m	0.92	ng	
30) Fluoranthene	19.17	202	54756	0.86	ng	# 87
33) Pyrene	19.53	202	33304	0.48	ng	# 91
35) Benzo(a)anthracene	21.27	228	4077m	0.05	ng	
38) Bis(2-ethylhexyl)phthalate	21.20	149	23767	0.74	ng	# 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE052516\
Data File : BE091866.D
Acq On : 25 May 2016 19:50
Operator : UM/SJ
Sample : H3065-04
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SS043MW010-160511

Manual Integrations
APPROVED

umangi
6/22/2016 5:40:08 PM

Quant Time: May 26 01:48:00 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE052516.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 25 19:42:30 2016
Response via : Initial Calibration

