

Data Path : Z:\HPCHEM1\BNA E\DATA\BE052516\
 Data File : BE091867.D
 Acq On : 25 May 2016 20:26
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 26 01:49:34 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE052516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 25 19:42:30 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	113	0.00
2	1,4-Dioxane	0.737	0.641	13.0	106	0.00
3	n-Nitrosodimethylamine	0.918	0.773	15.8	108	0.00
4 S	2-Fluorophenol	1.429	1.321	7.6	114	0.00
5 S	Phenol-d6	1.874	1.796	4.2	123	0.00
6	bis(2-Chloroethyl)ether	1.569	1.470	6.3	116	-0.02
7	n-Nitroso-di-n-propylamine	1.452	1.351	7.0	123	0.00
8 I	Naphthalene-d8	1.000	1.000	0.0	122	0.00
9 S	Nitrobenzene-d5	0.406	0.376	7.4	123	0.00
10	Nitrobenzene	0.394	0.361	8.4	128	0.00
11	bis(2-Chloroethoxy)methane	0.474	0.378	20.3	123	0.00
12	Naphthalene	1.080	1.043	3.4	125	0.00
13	Hexachlorobutadiene	0.166	0.159	4.2	121	0.00
14	2-Methylnaphthalene	0.721	0.706	2.1	129	0.00
15 I	Acenaphthene-d10	1.000	1.000	0.0	128	0.00
16 S	2,4,6-Tribromophenol	0.223	0.185	17.0	120	0.00
17 S	2-Fluorobiphenyl	1.487	1.408	5.3	126	0.00
18	Acenaphthylene	2.458	2.265	7.9	129	0.00
19	2,6-Dinitrotoluene	0.341	0.279	18.2	125	-0.02
20	Acenaphthene	1.337	1.287	3.7	130	0.00
21	2,4-Dinitrotoluene	0.393	0.301	23.4	130	0.00
22	Fluorene	1.755	1.658	5.5	129	0.00
23	4-Chlorophenyl-phenylether	0.883	0.808	8.5	123	0.00
24 I	Phenanthrene-d10	1.000	1.000	0.0	118	0.00
25	4-Bromophenyl-phenylether	0.179	0.173	3.4	122	0.00
26	Hexachlorobenzene	0.170	0.196	-15.3	140	0.00
27	Pentachlorophenol	0.066	0.065	1.5	150#	-0.02
28	Phenanthrene	1.090	1.013	7.1	118	0.00
29	Anthracene	0.999	0.953	4.6	123	0.00
30	Fluoranthene	1.273	1.147	9.9	116	-0.02
31 I	Chrysene-d12	1.000	1.000	0.0	95	0.00
32	Benzidine	0.400	0.391	2.3	116	0.00
33	Pyrene	1.061	1.177	-10.9	115	0.00
34 S	Terphenyl-d14	0.755	0.813	-7.7	109	0.00
35	Benzo(a)anthracene	1.333	1.441	-8.1	109	0.00
36	3,3'-Dichlorobenzidine	0.732	0.747	-2.0	107	0.00
37	Chrysene	1.222	1.279	-4.7	102	0.00
38	Bis(2-ethylhexyl)phthalate	0.484	0.592	-22.3	137	0.00
39	Indeno(1,2,3-cd)pyrene	1.250	1.314	-5.1	107	0.00
40 I	Perylene-d12	1.000	1.000	0.0	101	0.00
41	Benzo(b)fluoranthene	1.349	1.397	-3.6	119	0.00

Data Path : Z:\HPCHEM1\BNA E\DATA\BE052516\
Data File : BE091867.D
Acq On : 25 May 2016 20:26
Operator : UM/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 26 01:49:34 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE052516.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 25 19:42:30 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
42	Benzo(k)fluoranthene	1.225	1.130	7.8	91	0.00
43 C	Benzo(a)pyrene	1.223	1.213	0.8	107	0.00
44	Dibenzo(a,h)anthracene	1.172	1.161	0.9	106	0.00
45	Benzo(a,h,i)perylene	1.189	1.186	0.3	107	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0