

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052819\
 Data File : BE099757.D
 Acq On : 28 May 2019 15:14
 Operator : JU/SJ
 Sample : PB120100BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 PB120100BL

Quant Time: May 28 16:11:41 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 20 19:29:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.80	152	1263	0.40	ng	-0.01
7) Naphthalene-d8	10.61	136	3976	0.40	ng	0.00
13) Acenaphthene-d10	14.49	164	2731	0.40	ng	0.00
19) Phenanthrene-d10	17.23	188	8697	0.40	ng	0.00
27) Chrysene-d12	21.41	240	14128	0.40	ng	0.00
34) Perylene-d12	23.95	264	17392	0.40	ng	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.37	112	1212	0.38	ng	0.00
5) Phenol-d6	6.96	99	1545	0.37	ng	-0.01
8) Nitrobenzene-d5	8.98	82	1023	0.30	ng	0.00
11) 2-Methylnaphthalene-d10	12.22	152	2349	0.34	ng	0.00
14) 2,4,6-Tribromophenol	15.98	330	607	0.42	ng	0.00
15) 2-Fluorobiphenyl	13.12	172	3639	0.35	ng	0.01
25) Fluoranthene-d10	19.26	212	44130	0.37	ng	0.00
29) Terphenyl-d14	19.86	244	9501	0.38	ng	-0.01

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052819\
Data File : BE099757.D
Acq On : 28 May 2019 15:14
Operator : JU/SJ
Sample : PB120100BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
PB120100BL

Quant Time: May 28 16:11:41 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 20 19:29:40 2019
Response via : Initial Calibration

