

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052819\
 Data File : BE099755.D
 Acq On : 28 May 2019 13:47
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 28 15:10:30 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 20 19:29:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	72	0.00
2	1,4-Dioxane	0.592	0.564	4.7	72	-0.02
3	n-Nitrosodimethylamine	0.327	0.320	2.1	80	-0.02
4 S	2-Fluorophenol	1.011	0.923	8.7	75	0.00
5 S	Phenol-d6	1.338	1.243	7.1	81	0.00
6	bis(2-Chloroethyl)ether	1.190	1.096	7.9	76	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	75	0.00
8 S	Nitrobenzene-d5	0.346	0.306	11.6	77	-0.01
9	Naphthalene	4.278	3.862	9.7	76	-0.01
10	Hexachlorobutadiene	0.315	0.296	6.0	79	0.00
11 SURR	2-Methylnaphthalene-d10	0.690	0.628	9.0	77	-0.01
12	2-Methylnaphthalene	0.697	0.648	7.0	80	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	79	0.00
14 S	2,4,6-Tribromophenol	0.237	0.202	14.8	86	-0.01
15 S	2-Fluorobiphenyl	1.518	1.337	11.9	78	0.00
16	Acenaphthylene	1.864	1.731	7.1	85	-0.01
17	Acenaphthene	1.046	0.941	10.0	80	0.00
18	Fluorene	1.541	1.409	8.6	83	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	85	0.00
20	4-Bromophenyl-phenylether	0.231	0.196	15.2	83	-0.01
21	Hexachlorobenzene	0.236	0.206	12.7	84	0.00
22	Pentachlorophenol	0.068	0.066	2.9	142	-0.01
23	Phenanthrene	0.977	0.857	12.3	85	0.00
24	Anthracene	0.886	0.749	15.5	86	0.00
25 SURR	Fluoranthene-d10	5.466	4.781	12.5	85	0.00
26	Fluoranthene	1.545	1.370	11.3	86	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	88	0.00
28	Pvrene	1.005	0.997	0.8	88	-0.01
29 S	Terphenyl-d14	0.714	0.693	2.9	89	-0.01
30	Benzo(a)anthracene	1.137	1.168	-2.7	95	-0.01
31	Chrysene	1.215	1.167	4.0	85	0.00
32	Bis(2-ethylhexyl)phthalate	1.310	1.288	1.7	96	-0.01
33	Indeno(1,2,3-cd)pyrene	1.467	1.415	3.5	90	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	87	-0.01
35	Benzo(b)fluoranthene	1.096	1.003	8.5	89	-0.01
36	Benzo(k)fluoranthene	1.143	1.004	12.2	86	-0.01
37 C	Benzo(a)pyrene	1.055	0.950	10.0	89	-0.01
38	Dibenzo(a,h)anthracene	1.108	0.970	12.5	88	-0.02
39	Benzo(a,h,i)perylene	1.118	0.988	11.6	87	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052819\
Data File : BE099755.D
Acq On : 28 May 2019 13:47
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 28 15:10:30 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 20 19:29:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			