

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052819\
 Data File : BE099770.D
 Acq On : 28 May 2019 23:09
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 29 04:13:04 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 20 19:29:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	77	0.00
2	1,4-Dioxane	0.592	0.481	18.8	65	-0.02
3	n-Nitrosodimethylamine	0.327	0.294	10.1	79	-0.02
4 S	2-Fluorophenol	1.011	1.010	0.1	88	0.00
5 S	Phenol-d6	1.338	1.329	0.7	92	0.00
6	bis(2-Chloroethyl)ether	1.190	1.021	14.2	75	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	79	-0.01
8 S	Nitrobenzene-d5	0.346	0.314	9.2	84	-0.01
9	Naphthalene	4.278	3.815	10.8	79	-0.01
10	Hexachlorobutadiene	0.315	0.279	11.4	79	-0.01
11 SURR	2-Methylnaphthalene-d10	0.690	0.630	8.7	81	-0.01
12	2-Methylnaphthalene	0.697	0.641	8.0	83	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	87	-0.01
14 S	2,4,6-Tribromophenol	0.237	0.248	-4.6	117	-0.01
15 S	2-Fluorobiphenyl	1.518	1.273	16.1	82	0.00
16	Acenaphthylene	1.864	1.707	8.4	92	-0.01
17	Acenaphthene	1.046	0.942	9.9	88	0.00
18	Fluorene	1.541	1.432	7.1	92	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	91	0.00
20	4-Bromophenyl-phenylether	0.231	0.197	14.7	89	-0.01
21	Hexachlorobenzene	0.236	0.199	15.7	87	-0.01
22	Pentachlorophenol	0.068	0.056	17.6	129	-0.01
23	Phenanthrene	0.977	0.882	9.7	93	0.00
24	Anthracene	0.886	0.799	9.8	98	0.00
25 SURR	Fluoranthene-d10	5.466	4.897	10.4	93	0.00
26	Fluoranthene	1.545	1.391	10.0	93	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	92	-0.01
28	Pvrene	1.005	1.016	-1.1	95	-0.01
29 S	Terphenyl-d14	0.714	0.699	2.1	94	-0.01
30	Benzo(a)anthracene	1.137	1.145	-0.7	97	-0.01
31	Chrysene	1.215	1.194	1.7	92	0.00
32	Bis(2-ethylhexyl)phthalate	1.310	1.709	-30.5#	134	-0.01
33	Indeno(1,2,3-cd)pyrene	1.467	1.396	4.8	93	-0.03
34 I	Perylene-d12	1.000	1.000	0.0	91	-0.02
35	Benzo(b)fluoranthene	1.096	1.009	7.9	94	-0.01
36	Benzo(k)fluoranthene	1.143	0.990	13.4	89	-0.02
37 C	Benzo(a)pyrene	1.055	0.962	8.8	95	-0.02
38	Dibenzo(a,h)anthracene	1.108	0.967	12.7	92	-0.03
39	Benzo(a,h,i)perylene	1.118	0.994	11.1	92	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052819\
Data File : BE099770.D
Acq On : 28 May 2019 23:09
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 29 04:13:04 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 20 19:29:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			