

Data Path : Z:\HPCHEM1\BNA_E\Data\BE060817\
 Data File : BE093073.D
 Acq On : 8 Jun 2017 8:58
 Operator : SJ/MA
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.458

Quant Time: Jun 08 13:24:54 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 08 13:24:15 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	104	0.00
3	Naphthalene	0.992	0.983	0.9	105	0.00
4 SURR	2-Methylnaphthalene-d10	0.586	0.565	3.6	100	0.00
5	2-Methylnaphthalene	0.679	0.654	3.7	101	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	96	0.00
7	Acenaphthylene	1.717	1.669	2.8	93	0.00
8 C	Acenaphthene	1.385	1.393	-0.6	96	0.00
9	Fluorene	1.624	1.589	2.2	95	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	91	0.00
11	Pentachlorophenol	0.059	0.050	15.3	83	0.00
12	Phenanthrene	1.101	1.083	1.6	93	0.00
13	Anthracene	0.958	0.961	-0.3	93	0.00
14 SURR	Fluoranthene-d10	1.088	1.126	-3.5	96	0.00
15 C	Fluoranthene	1.350	1.392	-3.1	96	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	101	0.00
17	Pyrene	1.180	1.188	-0.7	97	0.00
18	Benzo(a)anthracene	1.067	1.090	-2.2	104	0.00
19	Chrysene	1.107	1.096	1.0	102	0.00
20 I	Perylene-d12	1.000	1.000	0.0	105	0.00
21	Benzo(b)fluoranthene	1.214	1.124	7.4	100	0.00
22	Benzo(k)fluoranthene	1.289	1.285	0.3	107	0.00
23 C	Benzo(a)pyrene	1.148	1.128	1.7	108	0.00
24	Indeno(1,2,3-cd)pyrene	1.329	1.298	2.3	106	0.00
25	Dibenzo(a,h)anthracene	1.119	1.104	1.3	106	0.00
26	Benzo(g,h,i)perylene	1.140	1.106	3.0	106	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0