

Data Path : Z:\HPCHEM1\BNA_E\Data\BE060817\
 Data File : BE093083.D
 Acq On : 8 Jun 2017 16:31
 Operator : SJ/MA
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.459

Quant Time: Jun 08 17:02:19 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 08 16:34:32 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	120	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	117	0.00
3	Naphthalene	0.992	0.981	1.1	117	0.00
4 SURR	2-Methylnaphthalene-d10	0.586	0.570	2.7	114	0.00
5	2-Methylnaphthalene	0.679	0.663	2.4	115	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	112	0.00
7	Acenaphthylene	1.717	1.553	9.6	101	0.00
8 C	Acenaphthene	1.385	1.359	1.9	109	0.00
9	Fluorene	1.624	1.574	3.1	110	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	108	0.00
11	Pentachlorophenol	0.059	0.043	27.1#	85	0.00
12	Phenanthrene	1.101	1.098	0.3	111	0.00
13	Anthracene	0.958	0.937	2.2	107	0.00
14 SURR	Fluoranthene-d10	1.088	1.085	0.3	109	0.00
15 C	Fluoranthene	1.350	1.338	0.9	108	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	111	0.00
17	Pyrene	1.180	1.239	-5.0	110	0.00
18	Benzo(a)anthracene	1.067	1.055	1.1	110	0.00
19	Chrysene	1.107	1.095	1.1	111	0.00
20 I	Perylene-d12	1.000	1.000	0.0	111	0.00
21	Benzo(b)fluoranthene	1.214	1.119	7.8	106	0.00
22	Benzo(k)fluoranthene	1.289	1.274	1.2	112	0.00
23 C	Benzo(a)pyrene	1.148	1.080	5.9	109	0.00
24	Indeno(1,2,3-cd)pyrene	1.329	1.272	4.3	110	0.00
25	Dibenzo(a,h)anthracene	1.119	1.084	3.1	110	0.00
26	Benzo(g,h,i)perylene	1.140	1.080	5.3	110	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0