

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE061217\
 Data File : BE093123.D
 Acq On : 13 Jun 2017 4:22
 Operator : SJ/MA
 Sample : I3460-01
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 MW-101

Manual Integrations
 APPROVED

mohammad
 6/13/2017 3:56:03 PM

Quant Time: Jun 13 05:20:43 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 12 14:07:54 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.75	152	702	0.40	ng	0.00
7) Naphthalene-d8	10.51	136	3560	0.40	ng	-0.02
13) Acenaphthene-d10	14.37	164	2123	0.40	ng	0.00
19) Phenanthrene-d10	17.13	188	4160	0.40	ng	0.00
27) Chrysene-d12	21.27	240	5384	0.40	ng	0.00
34) Perylene-d12	23.67	264	4548	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.35	112	490	0.24	ng	0.00
5) Phenol-d6	6.90	99	451	0.15	ng	0.00
8) Nitrobenzene-d5	8.88	82	1187	0.42	ng	0.00
11) 2-Methylnaphthalene-d10	12.11	152	2246	0.42	ng	0.00
14) 2,4,6-Tribromophenol	15.85	330	417	0.86	ng	0.00
15) 2-Fluorobiphenyl	12.99	172	4771	0.53	ng	0.00
25) Fluoranthene-d10	19.14	212	6212	0.45	ng	0.02
29) Terphenyl-d14	19.72	244	5650	0.54	ng	0.00
Target Compounds						
24) Anthracene	17.25	178	1215m	0.076	ng	Qvalue 94
32) Bis(2-ethylhexyl)phthalate	21.18	149	3777	0.633	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE061217\
Data File : BE093123.D
Acq On : 13 Jun 2017 4:22
Operator : SJ/MA
Sample : I3460-01
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
MW-101

Manual Integrations
APPROVED

mohammad
6/13/2017 3:56:03 PM

Quant Time: Jun 13 05:20:43 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 12 14:07:54 2017
Response via : Initial Calibration

