

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE061217\
 Data File : BE093125.D
 Acq On : 13 Jun 2017 5:34
 Operator : SJ/MA
 Sample : I3460-09
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 MW-118

Quant Time: Jun 13 07:58:44 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 12 14:07:54 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.75	152	600	0.40	ng	0.00
7) Naphthalene-d8	10.52	136	3054	0.40	ng	0.00
13) Acenaphthene-d10	14.37	164	2954	0.40	ng	0.00
19) Phenanthrene-d10	17.13	188	3862	0.40	ng	0.00
27) Chrysene-d12	21.27	240	5174	0.40	ng	0.00
34) Perylene-d12	23.67	264	4433	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.35	112	422	0.24	ng	0.00
5) Phenol-d6	6.90	99	348	0.13	ng	0.00
8) Nitrobenzene-d5	8.88	82	1142	0.47	ng	0.00
11) 2-Methylnaphthalene-d10	12.11	152	1972	0.43	ng	0.00
14) 2,4,6-Tribromophenol	15.85	330	353	0.52	ng	0.00
15) 2-Fluorobiphenyl	12.99	172	4533	0.36	ng	0.00
25) Fluoranthene-d10	19.14	212	5311	0.42	ng	0.02
29) Terphenyl-d14	19.72	244	4978	0.50	ng	0.00
Target Compounds						
28) Pyrene	19.51	202	10688	0.170	ng	# 95
32) Bis(2-ethylhexyl)phthalate	21.18	149	1099	0.178	ng	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE061217\
Data File : BE093125.D
Acq On : 13 Jun 2017 5:34
Operator : SJ/MA
Sample : I3460-09
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
MW-118

Quant Time: Jun 13 07:58:44 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 12 14:07:54 2017
Response via : Initial Calibration

