

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE061217\
 Data File : BE093102.D
 Acq On : 12 Jun 2017 13:27
 Operator : SJ/MA
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE061217

Quant Time: Jun 12 14:07:47 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 12 14:07:54 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
2	1,4-Dioxane	0.887	0.864	2.6	120	0.00
3	n-Nitrosodimethylamine	0.758	0.755	0.4	114	0.00
4 S	2-Fluorophenol	1.180	1.168	1.0	120	0.00
5 S	Phenol-d6	1.729	1.668	3.5	114	0.00
6	bis(2-Chloroethyl)ether	1.510	1.520	-0.7	117	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	105	-0.02
8 S	Nitrobenzene-d5	0.319	0.319	0.0	121	0.00
9	Naphthalene	1.073	1.057	1.5	117	0.00
10	Hexachlorobutadiene	0.146	0.141	3.4	118	0.00
11 SURR	2-Methylnaphthalene-d10	0.601	0.588	2.2	119	0.00
12	2-Methylnaphthalene	0.666	0.655	1.7	118	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	112	0.00
14 S	2,4,6-Tribromophenol	0.091	0.086	5.5	128	0.00
15 S	2-Fluorobiphenyl	1.708	1.660	2.8	121	0.00
16	Acenaphthylene	6.960	7.218	-3.7	139	0.00
17	Acenaphthene	1.297	1.310	-1.0	130	0.00
18	Fluorene	1.563	1.568	-0.3	129	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	113	0.00
20	4-Bromophenyl-phenylether	0.244	0.241	1.2	130	0.00
21	Hexachlorobenzene	0.214	0.183	14.5	125	0.00
22	Pentachlorophenol	0.069	0.059	14.5	121	0.00
23	Phenanthrene	1.808	1.697	6.1	125	0.00
24	Anthracene	1.532	1.448	5.5	138	0.00
25 SURR	Fluoranthene-d10	1.318	1.235	6.3	126	0.00
26	Fluoranthene	6.567	6.124	6.7	129	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
28	Pyrene	4.870	5.042	-3.5	112	0.00
29 S	Terphenyl-d14	0.772	0.827	-7.1	116	0.00
30	Benzo(a)anthracene	1.080	1.072	0.7	110	0.00
31	Chrysene	1.153	1.151	0.2	110	0.00
32	Bis(2-ethylhexyl)phthalate	0.525	0.486	7.4	115	0.00
33	Indeno(1,2,3-cd)pyrene	4.306	4.124	4.2	107	0.00
34 I	Perylene-d12	1.000	1.000	0.0	94	0.00
35	Benzo(b)fluoranthene	1.285	1.264	1.6	108	0.00
36	Benzo(k)fluoranthene	1.240	1.230	0.8	108	0.00
37 C	Benzo(a)pyrene	1.111	1.090	1.9	108	0.00
38	Dibenzo(a,h)anthracene	1.158	1.138	1.7	108	0.00
39	Benzo(g,h,i)perylene	4.678	4.495	3.9	106	0.02

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE061217\
Data File : BE093102.D
Acq On : 12 Jun 2017 13:27
Operator : SJ/MA
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE061217

Quant Time: Jun 12 14:07:47 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 12 14:07:54 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		