

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE061217\
 Data File : BE093104.D
 Acq On : 12 Jun 2017 16:00
 Operator : SJ/MA
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 13 01:10:35 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 12 14:07:54 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	104	0.00
2	1,4-Dioxane	0.400	0.354	11.5	110	0.00
3	n-Nitrosodimethylamine	0.400	0.357	10.8	102	0.00
4 S	2-Fluorophenol	0.400	0.325	18.8	99	0.00
5 S	Phenol-d6	0.400	0.318	20.5	95	0.00
6	bis(2-Chloroethyl)ether	0.400	0.347	13.3	102	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	101	0.00
8 S	Nitrobenzene-d5	0.400	0.349	12.8	102	0.00
9	Naphthalene	0.400	0.356	11.0	102	0.00
10	Hexachlorobutadiene	0.400	0.361	9.8	106	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.350	12.5	103	0.00
12	2-Methylnaphthalene	0.400	0.356	11.0	103	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	105	0.00
14 S	2,4,6-Tribromophenol	0.400	0.321	19.8	102	0.00
15 S	2-Fluorobiphenyl	0.400	0.360	10.0	105	0.00
16	Acenaphthylene	0.400	0.339	15.3	107	0.00
17	Acenaphthene	0.400	0.362	9.5	109	0.00
18	Fluorene	0.400	0.356	11.0	108	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	102	0.00
20	4-Bromophenyl-phenylether	0.400	0.356	11.0	105	0.00
21	Hexachlorobenzene	0.400	0.337	15.8	112	0.00
22	Pentachlorophenol	0.400	0.311	22.3	100	0.00
23	Phenanthrene	0.400	0.338	15.5	102	0.00
24	Anthracene	0.400	0.329	17.8	108	0.00
25 SURR	Fluoranthene-d10	0.400	0.337	15.8	102	0.00
26	Fluoranthene	0.400	0.339	15.3	106	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	101	0.00
28	Pyrene	0.400	0.376	6.0	104	0.00
29 S	Terphenyl-d14	0.400	0.369	7.8	102	0.00
30	Benzo(a)anthracene	0.400	0.343	14.2	97	0.00
31	Chrysene	0.400	0.344	14.0	97	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.325	18.8	88	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.336	16.0	96	0.00
34 I	Perylene-d12	0.400	0.400	0.0	98	0.00
35	Benzo(b)fluoranthene	0.400	0.351	12.3	100	0.00
36	Benzo(k)fluoranthene	0.400	0.345	13.8	98	0.00
37 C	Benzo(a)pyrene	0.400	0.330	17.5	94	0.00
38	Dibenzo(a,h)anthracene	0.400	0.345	13.8	99	0.00
39	Benzo(g,h,i)perylene	0.400	0.341	14.8	98	0.00

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE061217\
Data File : BE093104.D
Acq On : 12 Jun 2017 16:00
Operator : SJ/MA
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 13 01:10:35 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 12 14:07:54 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				