

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE061217\
 Data File : BE093104.D
 Acq On : 12 Jun 2017 16:00
 Operator : SJ/MA
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 13 01:10:35 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 12 14:07:54 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
2	1,4-Dioxane	0.887	0.785	11.5	110	0.00
3	n-Nitrosodimethylamine	0.758	0.676	10.8	102	0.00
4 S	2-Fluorophenol	1.180	0.960	18.6	99	0.00
5 S	Phenol-d6	1.729	1.375	20.5	95	0.00
6	bis(2-Chloroethyl)ether	1.510	1.309	13.3	102	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	101	0.00
8 S	Nitrobenzene-d5	0.319	0.278	12.9	102	0.00
9	Naphthalene	1.073	0.956	10.9	102	0.00
10	Hexachlorobutadiene	0.146	0.131	10.3	106	0.00
11 SURR	2-Methylnaphthalene-d10	0.601	0.526	12.5	103	0.00
12	2-Methylnaphthalene	0.666	0.593	11.0	103	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	105	0.00
14 S	2,4,6-Tribromophenol	0.091	0.073	19.8	102	0.00
15 S	2-Fluorobiphenyl	1.708	1.539	9.9	105	0.00
16	Acenaphthylene	6.960	5.901	15.2	107	0.00
17	Acenaphthene	1.297	1.173	9.6	109	0.00
18	Fluorene	1.563	1.393	10.9	108	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	102	0.00
20	4-Bromophenyl-phenylether	0.244	0.217	11.1	105	0.00
21	Hexachlorobenzene	0.214	0.181	15.4	112	0.00
22	Pentachlorophenol	0.069	0.054	21.7	100	0.00
23	Phenanthrene	1.808	1.526	15.6	102	0.00
24	Anthracene	1.532	1.258	17.9	108	0.00
25 SURR	Fluoranthene-d10	1.318	1.110	15.8	102	0.00
26	Fluoranthene	6.567	5.560	15.3	106	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	101	0.00
28	Pyrene	4.870	4.572	6.1	104	0.00
29 S	Terphenyl-d14	0.772	0.711	7.9	102	0.00
30	Benzo(a)anthracene	1.080	0.926	14.3	97	0.00
31	Chrysene	1.153	0.992	14.0	97	0.00
32	Bis(2-ethylhexyl)phthalate	0.525	0.362	31.0#	88	0.00
33	Indeno(1,2,3-cd)pyrene	4.306	3.612	16.1	96	0.00
34 I	Perylene-d12	1.000	1.000	0.0	98	0.00
35	Benzo(b)fluoranthene	1.285	1.128	12.2	100	0.00
36	Benzo(k)fluoranthene	1.240	1.071	13.6	98	0.00
37 C	Benzo(a)pyrene	1.111	0.918	17.4	94	0.00
38	Dibenzo(a,h)anthracene	1.158	0.998	13.8	99	0.00
39	Benzo(g,h,i)perylene	4.678	3.993	14.6	98	0.00

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE061217\
Data File : BE093104.D
Acq On : 12 Jun 2017 16:00
Operator : SJ/MA
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 13 01:10:35 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 12 14:07:54 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			