

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE061217\
 Data File : BE093120.D
 Acq On : 13 Jun 2017 2:29
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 13 04:59:33 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 12 14:07:54 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	78	0.00
2	1,4-Dioxane	0.400	0.441	-10.2	102	0.00
3	n-Nitrosodimethylamine	0.400	0.470	-17.5	100	0.00
4 S	2-Fluorophenol	0.400	0.514	-28.5#	117	0.00
5 S	Phenol-d6	0.400	0.488	-22.0	108	0.00
6	bis(2-Chloroethyl)ether	0.400	0.464	-16.0	101	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	83	-0.02
8 S	Nitrobenzene-d5	0.400	0.495	-23.7	118	0.00
9	Naphthalene	0.400	0.433	-8.2	101	0.00
10	Hexachlorobutadiene	0.400	0.429	-7.2	103	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.443	-10.7	106	0.00
12	2-Methylnaphthalene	0.400	0.444	-11.0	105	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	94	0.00
14 S	2,4,6-Tribromophenol	0.400	0.603	-50.7#	171	0.00
15 S	2-Fluorobiphenyl	0.400	0.407	-1.7	106	0.00
16	Acenaphthylene	0.400	0.437	-9.2	123	0.00
17	Acenaphthene	0.400	0.430	-7.5	116	0.00
18	Fluorene	0.400	0.443	-10.7	119	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	87	0.00
20	4-Bromophenyl-phenylether	0.400	0.439	-9.7	110	0.00
21	Hexachlorobenzene	0.400	0.450	-12.5	127	0.00
22	Pentachlorophenol	0.400	0.774	-93.5#	212	0.00
23	Phenanthrene	0.400	0.412	-3.0	105	0.00
24	Anthracene	0.400	0.451	-12.7	126	0.00
25 SURR	Fluoranthene-d10	0.400	0.463	-15.8	119	0.00
26	Fluoranthene	0.400	0.443	-10.7	118	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	87	0.00
28	Pyrene	0.400	0.446	-11.5	106	0.00
29 S	Terphenyl-d14	0.400	0.418	-4.5	100	0.00
30	Benzo(a)anthracene	0.400	0.453	-13.2	110	0.00
31	Chrysene	0.400	0.424	-6.0	103	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.718	-79.5#	166	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.426	-6.5	105	0.00
34 I	Perylene-d12	0.400	0.400	0.0	85	0.00
35	Benzo(b)fluoranthene	0.400	0.452	-13.0	112	0.00
36	Benzo(k)fluoranthene	0.400	0.404	-1.0	99	0.00
37 C	Benzo(a)pyrene	0.400	0.446	-11.5	111	0.00
38	Dibenzo(a,h)anthracene	0.400	0.410	-2.5	102	0.00
39	Benzo(g,h,i)perylene	0.400	0.418	-4.5	104	0.00

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE061217\
Data File : BE093120.D
Acq On : 13 Jun 2017 2:29
Operator : SJ/MA
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 13 04:59:33 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 12 14:07:54 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				