

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE061217\
 Data File : BE093120.D
 Acq On : 13 Jun 2017 2:29
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 13 04:59:33 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 12 14:07:54 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	78	0.00
2	1,4-Dioxane	0.887	0.979	-10.4	102	0.00
3	n-Nitrosodimethylamine	0.758	0.890	-17.4	100	0.00
4 S	2-Fluorophenol	1.180	1.516	-28.5#	117	0.00
5 S	Phenol-d6	1.729	2.110	-22.0	108	0.00
6	bis(2-Chloroethyl)ether	1.510	1.750	-15.9	101	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	83	-0.02
8 S	Nitrobenzene-d5	0.319	0.395	-23.8	118	0.00
9	Naphthalene	1.073	1.162	-8.3	101	0.00
10	Hexachlorobutadiene	0.146	0.156	-6.8	103	0.00
11 SURR	2-Methylnaphthalene-d10	0.601	0.665	-10.6	106	0.00
12	2-Methylnaphthalene	0.666	0.740	-11.1	105	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	94	0.00
14 S	2,4,6-Tribromophenol	0.091	0.137	-50.5#	171#	0.00
15 S	2-Fluorobiphenyl	1.708	1.739	-1.8	106	0.00
16	Acenaphthylene	6.960	7.610	-9.3	123	0.00
17	Acenaphthene	1.297	1.396	-7.6	116	0.00
18	Fluorene	1.563	1.730	-10.7	119	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	87	0.00
20	4-Bromophenyl-phenylether	0.244	0.267	-9.4	110	0.00
21	Hexachlorobenzene	0.214	0.241	-12.6	127	0.00
22	Pentachlorophenol	0.069	0.134	-94.2#	212#	0.00
23	Phenanthrene	1.808	1.864	-3.1	105	0.00
24	Anthracene	1.532	1.727	-12.7	126	0.00
25 SURR	Fluoranthene-d10	1.318	1.524	-15.6	119	0.00
26	Fluoranthene	6.567	7.272	-10.7	118	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	87	0.00
28	Pyrene	4.870	5.436	-11.6	106	0.00
29 S	Terphenyl-d14	0.772	0.806	-4.4	100	0.00
30	Benzo(a)anthracene	1.080	1.223	-13.2	110	0.00
31	Chrysene	1.153	1.223	-6.1	103	0.00
32	Bis(2-ethylhexyl)phthalate	0.525	0.801	-52.6#	166#	0.00
33	Indeno(1,2,3-cd)pyrene	4.306	4.583	-6.4	105	0.00
34 I	Perylene-d12	1.000	1.000	0.0	85	0.00
35	Benzo(b)fluoranthene	1.285	1.451	-12.9	112	0.00
36	Benzo(k)fluoranthene	1.240	1.251	-0.9	99	0.00
37 C	Benzo(a)pyrene	1.111	1.238	-11.4	111	0.00
38	Dibenzo(a,h)anthracene	1.158	1.187	-2.5	102	0.00
39	Benzo(g,h,i)perylene	4.678	4.892	-4.6	104	0.00

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE061217\
Data File : BE093120.D
Acq On : 13 Jun 2017 2:29
Operator : SJ/MA
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 13 04:59:33 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 12 14:07:54 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		