

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE061217\
 Data File : BE093122.D
 Acq On : 13 Jun 2017 3:44
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 13 04:59:49 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 12 14:07:54 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	85	0.00
2	1,4-Dioxane	0.400	0.441	-10.2	111	0.00
3	n-Nitrosodimethylamine	0.400	0.470	-17.5	110	0.00
4 S	2-Fluorophenol	0.400	0.506	-26.5#	126	0.00
5 S	Phenol-d6	0.400	0.477	-19.2	115	0.00
6	bis(2-Chloroethyl)ether	0.400	0.458	-14.5	109	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	89	-0.02
8 S	Nitrobenzene-d5	0.400	0.470	-17.5	121	0.00
9	Naphthalene	0.400	0.427	-6.7	107	0.00
10	Hexachlorobutadiene	0.400	0.425	-6.2	110	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.440	-10.0	114	0.00
12	2-Methylnaphthalene	0.400	0.437	-9.2	111	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	101	0.00
14 S	2,4,6-Tribromophenol	0.400	0.564	-41.0#	171	0.00
15 S	2-Fluorobiphenyl	0.400	0.413	-3.2	115	0.00
16	Acenaphthylene	0.400	0.422	-5.5	127	0.00
17	Acenaphthene	0.400	0.430	-7.5	124	0.00
18	Fluorene	0.400	0.435	-8.7	126	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	93	0.00
20	4-Bromophenyl-phenylether	0.400	0.463	-15.8	124	0.00
21	Hexachlorobenzene	0.400	0.426	-6.5	128	0.00
22	Pentachlorophenol	0.400	0.635	-58.7#	186	0.00
23	Phenanthrene	0.400	0.430	-7.5	118	0.00
24	Anthracene	0.400	0.413	-3.2	123	0.00
25 SURR	Fluoranthene-d10	0.400	0.434	-8.5	119	0.00
26	Fluoranthene	0.400	0.414	-3.5	118	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	85	0.00
28	Pyrene	0.400	0.453	-13.2	106	0.00
29 S	Terphenyl-d14	0.400	0.449	-12.2	106	0.00
30	Benzo(a)anthracene	0.400	0.448	-12.0	107	0.00
31	Chrysene	0.400	0.428	-7.0	102	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.583	-45.7#	132	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.415	-3.7	101	0.00
34 I	Perylene-d12	0.400	0.400	0.0	81	0.00
35	Benzo(b)fluoranthene	0.400	0.451	-12.7	107	0.00
36	Benzo(k)fluoranthene	0.400	0.414	-3.5	97	0.00
37 C	Benzo(a)pyrene	0.400	0.442	-10.5	105	0.00
38	Dibenzo(a,h)anthracene	0.400	0.423	-5.7	100	0.00
39	Benzo(g,h,i)perylene	0.400	0.427	-6.7	101	0.00

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE061217\
Data File : BE093122.D
Acq On : 13 Jun 2017 3:44
Operator : SJ/MA
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 13 04:59:49 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 12 14:07:54 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				