

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE061217\
 Data File : BE093122.D
 Acq On : 13 Jun 2017 3:44
 Operator : SJ/MA
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Jun 13 04:59:49 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 12 14:07:54 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	85	0.00
2	1,4-Dioxane	0.887	0.979	-10.4	111	0.00
3	n-Nitrosodimethylamine	0.758	0.891	-17.5	110	0.00
4 S	2-Fluorophenol	1.180	1.494	-26.6#	126	0.00
5 S	Phenol-d6	1.729	2.060	-19.1	115	0.00
6	bis(2-Chloroethyl)ether	1.510	1.730	-14.6	109	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	89	-0.02
8 S	Nitrobenzene-d5	0.319	0.376	-17.9	121	0.00
9	Naphthalene	1.073	1.145	-6.7	107	0.00
10	Hexachlorobutadiene	0.146	0.155	-6.2	110	0.00
11 SURR	2-Methylnaphthalene-d10	0.601	0.661	-10.0	114	0.00
12	2-Methylnaphthalene	0.666	0.728	-9.3	111	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
14 S	2,4,6-Tribromophenol	0.091	0.129	-41.8#	171#	0.00
15 S	2-Fluorobiphenyl	1.708	1.765	-3.3	115	0.00
16	Acenaphthylene	6.960	7.339	-5.4	127	0.00
17	Acenaphthene	1.297	1.396	-7.6	124	0.00
18	Fluorene	1.563	1.700	-8.8	126	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	93	0.00
20	4-Bromophenyl-phenylether	0.244	0.282	-15.6	124	0.00
21	Hexachlorobenzene	0.214	0.228	-6.5	128	0.00
22	Pentachlorophenol	0.069	0.110	-59.4#	186#	0.00
23	Phenanthrene	1.808	1.944	-7.5	118	0.00
24	Anthracene	1.532	1.581	-3.2	123	0.00
25 SURR	Fluoranthene-d10	1.318	1.430	-8.5	119	0.00
26	Fluoranthene	6.567	6.799	-3.5	118	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	85	0.00
28	Pyrene	4.870	5.514	-13.2	106	0.00
29 S	Terphenyl-d14	0.772	0.867	-12.3	106	0.00
30	Benzo(a)anthracene	1.080	1.210	-12.0	107	0.00
31	Chrysene	1.153	1.235	-7.1	102	0.00
32	Bis(2-ethylhexyl)phthalate	0.525	0.644	-22.7	132	0.00
33	Indeno(1,2,3-cd)pyrene	4.306	4.470	-3.8	101	0.00
34 I	Perylene-d12	1.000	1.000	0.0	81	0.00
35	Benzo(b)fluoranthene	1.285	1.451	-12.9	107	0.00
36	Benzo(k)fluoranthene	1.240	1.283	-3.5	97	0.00
37 C	Benzo(a)pyrene	1.111	1.226	-10.4	105	0.00
38	Dibenzo(a,h)anthracene	1.158	1.226	-5.9	100	0.00
39	Benzo(g,h,i)perylene	4.678	4.998	-6.8	101	0.00

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE061217\
Data File : BE093122.D
Acq On : 13 Jun 2017 3:44
Operator : SJ/MA
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 13 04:59:49 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 12 14:07:54 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		