

Data File : BE093147.D
Acq On : 14 Jun 2017 1:13
Operator : SJ/MA
Sample : I3558-06
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
F5A60

Quant Time: Jun 14 08:39:45 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 14 08:39:04 2017
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.74	152	4564	0.40	ng/ul	0.00
2) Naphthalene-d8	10.52	136	22643	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.37	164	13952	0.40	ng/ul	0.00
10) Phenanthrene-d10	0.00	188	0	0.00	ng/ul	-17.10
16) Chrysene-d12	21.26	240	34383	0.40	ng/ul	0.00
20) Perylene-d12	23.51	264	1844602	0.40	ng/ul	-0.15
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.11	152	9347	0.28	ng/ul	0.00
14) Fluoranthene-d10	19.12	212	25641	0.00	ng/ul	0.00
Target Compounds						
17) Pyrene	19.49	202	4934	0.05	ng/ul#	1
19) Chrysene	21.29	228	1152	0.01	ng/ul#	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BE093147.D

Acq On : 14 Jun 2017 1:13

Operator : SJ/MA

Sample : I3558-06

Misc :

ALS Vial : 20 Sample Multiplier: 1

Instrument :

BNA_E

ClientSampleId :

F5A60


Quant Time: Jun 14 08:39:45 2017


Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Wed Jun 14 08:39:04 2017

Response via : Initial Calibration









#17
Pyrene
Concen: 0.05 ng/ul
RT: 19.49 min Scan# 2792
Delta R.T. -0.02 min
Lab File: BE093147.D
Acq: 14 Jun 2017 1:13

Instrument :
BNA_E
ClientSampleId :
F5A60

Tgt Ion:202 Resp: 4934
Ion Ratio Lower Upper
202 100
101 181.8 18.2 27.4#
100 1011.8 15.6 23.4#



#19
Chrysene
Concen: 0.01 ng/ul
RT: 21.29 min Scan# 3184
Delta R.T. -0.00 min
Lab File: BE093147.D
Acq: 14 Jun 2017 1:13

Tgt Ion:228 Resp: 1152
Ion Ratio Lower Upper
228 100
226 41.8 23.4 35.0#
229 38.0 17.7 26.5#

