

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE061317\
 Data File : BE093140.D
 Acq On : 13 Jun 2017 20:53
 Operator : SJ/MA
 Sample : I3522-20
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 F5D36

Manual Integrations
 APPROVED

mohammad
 6/14/2017 2:59:16 PM

Quant Time: Jun 14 08:07:15 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 14 01:25:15 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.74	152	4543	0.40	ng/ul	0.00
2) Naphthalene-d8	10.52	136	22060	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.37	164	13234	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.10	188	32481	0.40	ng/ul	0.00
16) Chrysene-d12	21.26	240	35520	0.40	ng/ul	0.00
20) Perylene-d12	23.66	264	38920	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.11	152	5678	0.18	ng/ul	0.00
14) Fluoranthene-d10	19.13	212	16181	0.18	ng/ul	0.00
Target Compounds				Qvalue		
3) Naphthalene	10.57	128	9649	0.176	ng/ul#	93
5) 2-Methylnaphthalene	12.19	142	1412	0.038	ng/ul	100
7) Acenaphthylene	14.08	152	9885	0.174	ng/ul	98
8) Acenaphthene	14.43	153	1605	0.035	ng/ul	97
9) Fluorene	15.42	166	2311	0.043	ng/ul	94
12) Phenanthrene	17.14	178	17860	0.200	ng/ul#	86
13) Anthracene	17.22	178	16437	0.211	ng/ul#	88
15) Fluoranthene	19.15	202	151751	1.385	ng/ul	84
17) Pyrene	19.51	202	164194	1.567	ng/ul#	87
18) Benzo(a)anthracene	21.24	228	116722	1.232	ng/ul#	85
19) Chrysene	21.29	228	159583	1.624	ng/ul	97
21) Benzo(b)fluoranthene	22.93	252	246395m	2.086	ng/ul	
22) Benzo(k)fluoranthene	22.97	252	88250m	0.704	ng/ul	
23) Benzo(a)pyrene	23.55	252	123764	1.108	ng/ul#	81
24) Indeno(1,2,3-cd)pyrene	26.17	276	73120	0.565	ng/ul#	87
25) Dibenzo(a,h)anthracene	26.19	278	19673	0.181	ng/ul#	90
26) Benzo(g,h,i)perylene	26.95	276	63228	0.570	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE061317\
Data File : BE093140.D
Acq On : 13 Jun 2017 20:53
Operator : SJ/MA
Sample : I3522-20
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
F5D36

Manual Integrations
APPROVED

mohammad
6/14/2017 2:59:16 PM

Quant Time: Jun 14 08:07:15 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 14 01:25:15 2017
Response via : Initial Calibration

