

Data Path : Z:\HPCHEM1\BNA_E\Data\BE061317\
 Data File : BE093162.D
 Acq On : 14 Jun 2017 10:28
 Operator : SJ/MA
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.403

Quant Time: Jun 14 10:58:51 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 08 17:42:17 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	86	0.00
3	Naphthalene	0.992	1.022	-3.0	90	0.00
4 SURR	2-Methylnaphthalene-d10	0.586	0.629	-7.3	92	0.00
5	2-Methylnaphthalene	0.679	0.723	-6.5	92	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	91	0.00
7	Acenaphthylene	1.717	1.782	-3.8	94	0.00
8 C	Acenaphthene	1.385	1.419	-2.5	92	0.00
9	Fluorene	1.624	1.637	-0.8	92	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	83	0.00
11	Pentachlorophenol	0.059	0.067	-13.6	100	0.00
12	Phenanthrene	1.101	1.151	-4.5	89	0.00
13	Anthracene	0.958	1.018	-6.3	89	0.00
14 SURR	Fluoranthene-d10	1.088	1.103	-1.4	85	0.00
15 C	Fluoranthene	1.350	1.346	0.3	84	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	74	0.00
17	Pyrene	1.180	1.346	-14.1	80	0.00
18	Benzo(a)anthracene	1.067	1.169	-9.6	81	0.00
19	Chrysene	1.107	1.151	-4.0	78	0.00
20 I	Perylene-d12	1.000	1.000	0.0	72	0.00
21	Benzo(b)fluoranthene	1.214	1.251	-3.0	76	0.00
22	Benzo(k)fluoranthene	1.289	1.300	-0.9	74	0.00
23 C	Benzo(a)pyrene	1.148	1.222	-6.4	79	0.00
24	Indeno(1,2,3-cd)pyrene	1.329	1.359	-2.3	76	0.00
25	Dibenzo(a,h)anthracene	1.119	1.133	-1.3	74	0.00
26	Benzo(g,h,i)perylene	1.140	1.146	-0.5	75	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0