

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE061417\
 Data File : BE093171.D
 Acq On : 14 Jun 2017 16:02
 Operator : SJ/MA
 Sample : I3522-17
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 F5B23

Manual Integrations
 APPROVED

mohammad
 6/15/2017 3:53:07 PM

Quant Time: Jun 15 03:29:49 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 15 02:27:21 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.74	152	3310	0.40	ng/ul	0.00
2) Naphthalene-d8	10.52	136	16679	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.37	164	10439	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.10	188	24134m	0.40	ng/ul	0.00
16) Chrysene-d12	21.26	240	26124	0.40	ng/ul	0.00
20) Perylene-d12	23.66	264	19913m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.11	152	6889	0.28	ng/ul	0.00
14) Fluoranthene-d10	19.12	212	18702	0.28	ng/ul	0.00
Target Compounds				Qvalue		
3) Naphthalene	10.57	128	36038	0.871	ng/ul#	91
5) 2-Methylnaphthalene	12.18	142	5877	0.208	ng/ul	98
7) Acenaphthylene	14.09	152	37910	0.846	ng/ul	95
8) Acenaphthene	14.43	153	5682	0.157	ng/ul	99
9) Fluorene	15.42	166	4606	0.109	ng/ul	91
12) Phenanthrene	17.14	178	61083	0.920	ng/ul#	87
13) Anthracene	17.22	178	55306	0.957	ng/ul#	87
15) Fluoranthene	19.15	202	432357	5.309	ng/ul	83
17) Pyrene	19.51	202	490711	6.369	ng/ul#	86
18) Benzo(a)anthracene	21.24	228	327475	4.699	ng/ul#	86
19) Chrysene	21.29	228	435378	6.023	ng/ul	98
21) Benzo(b)fluoranthene	22.93	252	721047	11.932	ng/ul	87
22) Benzo(k)fluoranthene	22.97	252	244857m	3.815	ng/ul	
23) Benzo(a)pyrene	23.56	252	400745m	7.013	ng/ul	
24) Indeno(1,2,3-cd)pyrene	26.18	276	242336m	3.663	ng/ul	
25) Dibenzo(a,h)anthracene	26.19	278	67643m	1.214	ng/ul	
26) Benzo(g,h,i)perylene	26.95	276	197079	3.473	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE061417\
Data File : BE093171.D
Acq On : 14 Jun 2017 16:02
Operator : SJ/MA
Sample : I3522-17
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
F5B23

Manual Integrations
APPROVED

mohammad
6/15/2017 3:53:07 PM

Quant Time: Jun 15 03:29:49 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 15 02:27:21 2017
Response via : Initial Calibration

