

Data Path : Z:\HPCHEM1\BNA_E\Data\BE061417\
 Data File : BE093164.D
 Acq On : 14 Jun 2017 11:42
 Operator : SJ/MA
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.404

Quant Time: Jun 15 02:25:00 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 14 08:39:04 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	81	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	77	0.00
3	Naphthalene	0.992	1.028	-3.6	81	0.00
4 SURR	2-Methylnaphthalene-d10	0.586	0.617	-5.3	81	0.00
5	2-Methylnaphthalene	0.679	0.713	-5.0	82	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	79	0.00
7	Acenaphthylene	1.717	1.744	-1.6	80	0.00
8 C	Acenaphthene	1.385	1.421	-2.6	81	0.00
9	Fluorene	1.624	1.631	-0.4	80	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	75	0.00
11	Pentachlorophenol	0.059	0.069	-16.9	94	0.00
12	Phenanthrene	1.101	1.132	-2.8	79	0.00
13	Anthracene	0.958	1.019	-6.4	81	0.00
14 SURR	Fluoranthene-d10	1.088	1.107	-1.7	78	0.00
15 C	Fluoranthene	1.350	1.355	-0.4	76	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	70	0.00
17	Pyrene	1.180	1.277	-8.2	72	0.00
18	Benzo(a)anthracene	1.067	1.148	-7.6	76	0.00
19	Chrysene	1.107	1.149	-3.8	74	0.00
20 I	Perylene-d12	1.000	1.000	0.0	68	0.00
21	Benzo(b)fluoranthene	1.214	1.286	-5.9	74	0.00
22	Benzo(k)fluoranthene	1.289	1.283	0.5	69	0.00
23 C	Benzo(a)pyrene	1.148	1.221	-6.4	76	0.00
24	Indeno(1,2,3-cd)pyrene	1.329	1.369	-3.0	73	-0.01
25	Dibenzo(a,h)anthracene	1.119	1.129	-0.9	70	-0.01
26	Benzo(g,h,i)perylene	1.140	1.151	-1.0	72	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0