

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE061417\
 Data File : BE093178.D
 Acq On : 14 Jun 2017 20:24
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.405

Quant Time: Jun 15 02:29:07 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 15 02:27:21 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	80	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	82	0.00
3	Naphthalene	0.992	1.019	-2.7	86	0.00
4 SURR	2-Methylnaphthalene-d10	0.586	0.642	-9.6	90	0.00
5	2-Methylnaphthalene	0.679	0.743	-9.4	90	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	92	0.00
7	Acenaphthylene	1.717	1.925	-12.1	103	0.00
8 C	Acenaphthene	1.385	1.425	-2.9	94	0.00
9	Fluorene	1.624	1.660	-2.2	95	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	84	0.00
11	Pentachlorophenol	0.059	0.081	-37.3#	124	0.00
12	Phenanthrene	1.101	1.133	-2.9	89	0.00
13	Anthracene	0.958	1.085	-13.3	97	0.00
14 SURR	Fluoranthene-d10	1.088	1.158	-6.4	91	0.00
15 C	Fluoranthene	1.350	1.375	-1.9	87	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	74	0.00
17	Pyrene	1.180	1.380	-16.9	82	0.00
18	Benzo(a)anthracene	1.067	1.215	-13.9	84	0.00
19	Chrysene	1.107	1.141	-3.1	77	0.00
20 I	Perylene-d12	1.000	1.000	0.0	73	0.00
21	Benzo(b)fluoranthene	1.214	1.255	-3.4	78	0.00
22	Benzo(k)fluoranthene	1.289	1.278	0.9	74	0.00
23 C	Benzo(a)pyrene	1.148	1.236	-7.7	83	0.00
24	Indeno(1,2,3-cd)pyrene	1.329	1.365	-2.7	78	0.00
25	Dibenzo(a,h)anthracene	1.119	1.143	-2.1	77	0.00
26	Benzo(g,h,i)perylene	1.140	1.135	0.4	76	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0