

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE061417\
 Data File : BE093196.D
 Acq On : 15 Jun 2017 7:27
 Operator : SJ/MA
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.406

Quant Time: Jun 15 07:57:03 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 15 03:59:13 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	65	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	72	0.00
3	Naphthalene	0.992	1.024	-3.2	75	0.00
4 SURR	2-Methylnaphthalene-d10	0.586	0.646	-10.2	79	0.00
5	2-Methylnaphthalene	0.679	0.752	-10.8	80	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	79	0.00
7	Acenaphthylene	1.717	1.928	-12.3	89	0.00
8 C	Acenaphthene	1.385	1.427	-3.0	81	0.00
9	Fluorene	1.624	1.620	0.2	80	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	70	0.00
11	Pentachlorophenol	0.059	0.073	-23.7#	93	0.00
12	Phenanthrene	1.101	1.116	-1.4	73	0.00
13	Anthracene	0.958	1.074	-12.1	80	0.00
14 SURR	Fluoranthene-d10	1.088	1.099	-1.0	72	0.00
15 C	Fluoranthene	1.350	1.318	2.4	70	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	60	0.00
17	Pyrene	1.180	1.379	-16.9	66	0.00
18	Benzo(a)anthracene	1.067	1.217	-14.1	68	0.00
19	Chrysene	1.107	1.136	-2.6	62	0.00
20 I	Perylene-d12	1.000	1.000	0.0	60	0.00
21	Benzo(b)fluoranthene	1.214	1.307	-7.7	67	0.00
22	Benzo(k)fluoranthene	1.289	1.238	4.0	59	0.00
23 C	Benzo(a)pyrene	1.148	1.224	-6.6	67	0.00
24	Indeno(1,2,3-cd)pyrene	1.329	1.378	-3.7	65	0.00
25	Dibenzo(a,h)anthracene	1.119	1.143	-2.1	63	0.00
26	Benzo(g,h,i)perylene	1.140	1.156	-1.4	64	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0