

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE061417\
 Data File : BE093205.D
 Acq On : 15 Jun 2017 12:58
 Operator : SJ/MA
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.407

Quant Time: Jun 15 13:54:05 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 15 07:58:47 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	59	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	65	0.00
3	Naphthalene	0.992	1.020	-2.8	68	0.00
4 SURR	2-Methylnaphthalene-d10	0.586	0.645	-10.1	72	0.00
5	2-Methylnaphthalene	0.679	0.747	-10.0	72	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	70	0.00
7	Acenaphthylene	1.717	2.056	-19.7	84	0.00
8 C	Acenaphthene	1.385	1.434	-3.5	73	0.00
9	Fluorene	1.624	1.660	-2.2	73	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	60	0.00
11	Pentachlorophenol	0.059	0.087	-47.5#	95	0.00
12	Phenanthrene	1.101	1.124	-2.1	63	0.00
13	Anthracene	0.958	1.145	-19.5	73	0.00
14 SURR	Fluoranthene-d10	1.088	1.149	-5.6	65	0.00
15 C	Fluoranthene	1.350	1.355	-0.4	61	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	49	0.00
17	Pyrene	1.180	1.451	-23.0#	57	0.00
18	Benzo(a)anthracene	1.067	1.250	-17.2	58	0.00
19	Chrysene	1.107	1.149	-3.8	52	0.00
20 I	Perylene-d12	1.000	1.000	0.0	52	0.00
21	Benzo(b)fluoranthene	1.214	1.227	-1.1	54	0.00
22	Benzo(k)fluoranthene	1.289	1.242	3.6	51	0.00
23 C	Benzo(a)pyrene	1.148	1.190	-3.7	56	0.00
24	Indeno(1,2,3-cd)pyrene	1.329	1.381	-3.9	56	0.00
25	Dibenzo(a,h)anthracene	1.119	1.132	-1.2	54	0.00
26	Benzo(g,h,i)perylene	1.140	1.138	0.2	54	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0