

Data Path : Z:\HPCHEM1\BNA_E\Data\BE061517\
 Data File : BE093237.D
 Acq On : 16 Jun 2017 13:47
 Operator : SJ/MA
 Sample : I3646-02RE
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 B0AD1RE

Quant Time: Jun 16 14:42:08 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 16 05:36:31 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.74	152	3402	0.40	ng/ul	0.00
2) Naphthalene-d8	10.52	136	15700	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.37	164	9122	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.12	188	33558	0.40	ng/ul	0.02
16) Chrysene-d12	21.36	240	12506	0.40	ng/ul	0.10
20) Perylene-d12	23.64	264	540560	0.40	ng/ul	-0.03
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.11	152	5714	0.25	ng/ul	0.00
14) Fluoranthene-d10	19.16	212	5545	0.06	ng/ul	0.04
Target Compounds				Qvalue		
3) Naphthalene	10.57	128	15996	0.411	ng/ul#	94
5) 2-Methylnaphthalene	12.19	142	6044	0.227	ng/ul	98
7) Acenaphthylene	13.94	152	17911	0.458	ng/ul	98
8) Acenaphthene	14.46	153	6757	0.214	ng/ul#	47
9) Fluorene	15.42	166	4209	0.114	ng/ul#	74
12) Phenanthrene	17.15	178	17323	0.188	ng/ul#	27
13) Anthracene	17.27	178	22063	0.274	ng/ul#	1
15) Fluoranthene	19.19	202	20295	0.179	ng/ul	80
17) Pyrene	19.50	202	7267	0.197	ng/ul	97
18) Benzo(a)anthracene	21.33	228	28886	0.866	ng/ul#	1
19) Chrysene	21.33	228	28886	0.835	ng/ul#	1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\Data\BE061517\
Data File : BE093237.D
Acq On : 16 Jun 2017 13:47
Operator : SJ/MA
Sample : I3646-02RE
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
B0AD1RE

Quant Time: Jun 16 14:42:08 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 16 05:36:31 2017
Response via : Initial Calibration

