

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE061517\
 Data File : BE093215.D
 Acq On : 15 Jun 2017 19:17
 Operator : SJ/MA
 Sample : I3522-05DL 5X
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 F5D30DL

Manual Integrations
 APPROVED

mohammad
 6/19/2017 11:06:40 AM

Quant Time: Jun 16 05:08:37 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 16 04:43:11 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.74	152	2531	0.40	ng/ul	0.00
2) Naphthalene-d8	10.52	136	12878	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.37	164	7556	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.10	188	18540	0.40	ng/ul	0.00
16) Chrysene-d12	21.26	240	20858	0.40	ng/ul	0.00
20) Perylene-d12	23.67	264	16855	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.11	152	990	0.05	ng/ul	0.00
14) Fluoranthene-d10	19.13	212	2625	0.05	ng/ul	0.00
Target Compounds				Qvalue		
3) Naphthalene	10.57	128	1794	0.056	ng/ul#	92
5) 2-Methylnaphthalene	12.18	142	459	0.021	ng/ul	100
7) Acenaphthylene	14.09	152	35331	1.090	ng/ul	96
8) Acenaphthene	14.43	153	879	0.034	ng/ul#	94
9) Fluorene	15.42	166	1234	0.040	ng/ul#	93
12) Phenanthrene	17.14	178	22530	0.442	ng/ul#	80
13) Anthracene	17.22	178	50015	1.126	ng/ul#	86
15) Fluoranthene	19.16	202	447430	7.152	ng/ul	83
17) Pyrene	19.52	202	499056	8.113	ng/ul#	87
18) Benzo(a)anthracene	21.24	228	238401	4.285	ng/ul#	85
19) Chrysene	21.30	228	328477	5.691	ng/ul	99
21) Benzo(b)fluoranthene	22.93	252	528596m	10.334	ng/ul	
22) Benzo(k)fluoranthene	22.98	252	191418m	3.524	ng/ul	
23) Benzo(a)pyrene	23.56	252	223700	4.625	ng/ul#	81
24) Indeno(1,2,3-cd)pyrene	26.19	276	137355	2.453	ng/ul#	86
25) Dibenzo(a,h)anthracene	26.20	278	35697	0.757	ng/ul	91
26) Benzo(g,h,i)perylene	26.97	276	98954	2.060	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE061517\
Data File : BE093215.D
Acq On : 15 Jun 2017 19:17
Operator : SJ/MA
Sample : I3522-05DL 5X
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
F5D30DL

Manual Integrations
APPROVED

mohammad
6/19/2017 11:06:40 AM

Quant Time: Jun 16 05:08:37 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 16 04:43:11 2017
Response via : Initial Calibration

