

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE061517\
 Data File : BE093220.D
 Acq On : 15 Jun 2017 22:22
 Operator : SJ/MA
 Sample : I3521-18DL 5X
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 F5D54DL

Manual Integrations
 APPROVED

mohammad
 6/19/2017 11:06:53 AM

Quant Time: Jun 16 05:27:32 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 16 04:43:11 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.74	152	2687	0.40	ng/ul	0.00
2) Naphthalene-d8	10.52	136	13944	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.37	164	8150	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.10	188	19451	0.40	ng/ul	0.00
16) Chrysene-d12	21.26	240	20546	0.40	ng/ul	0.00
20) Perylene-d12	23.67	264	19638	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.11	152	1486	0.07	ng/ul	0.00
14) Fluoranthene-d10	19.13	212	4007	0.08	ng/ul	0.00
Target Compounds				Qvalue		
3) Naphthalene	10.57	128	1650	0.048	ng/ul#	90
5) 2-Methylnaphthalene	12.18	142	463	0.020	ng/ul	97
7) Acenaphthylene	14.09	152	6886	0.197	ng/ul	98
12) Phenanthrene	17.13	178	6535	0.122	ng/ul#	87
13) Anthracene	17.22	178	8571	0.184	ng/ul#	87
15) Fluoranthene	19.16	202	54715	0.834	ng/ul	83
17) Pyrene	19.52	202	74047	1.222	ng/ul#	84
18) Benzo(a)anthracene	21.24	228	38119	0.696	ng/ul#	85
19) Chrysene	21.29	228	42097	0.740	ng/ul	98
21) Benzo(b)fluoranthene	22.93	252	82478m	1.384	ng/ul	
22) Benzo(k)fluoranthene	22.98	252	28649m	0.453	ng/ul	
23) Benzo(a)pyrene	23.56	252	42718	0.758	ng/ul#	82
24) Indeno(1,2,3-cd)pyrene	26.18	276	29668	0.455	ng/ul#	87
25) Dibenzo(a,h)anthracene	26.20	278	7304	0.133	ng/ul#	85
26) Benzo(g,h,i)perylene	26.96	276	24625	0.440	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE061517\
Data File : BE093220.D
Acq On : 15 Jun 2017 22:22
Operator : SJ/MA
Sample : I3521-18DL 5X
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
F5D54DL

Manual Integrations
APPROVED

mohammad
6/19/2017 11:06:53 AM

Quant Time: Jun 16 05:27:32 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 16 04:43:11 2017
Response via : Initial Calibration

