

Data File : BE093238.D
Acq On : 16 Jun 2017 14:25
Operator : SJ/MA
Sample : I3521-06DL 5X
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
F5C54DL

Manual Integrations
APPROVED

mohammad
6/19/2017 11:07:46 AM

Quant Time: Jun 19 07:30:18 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 16 05:36:31 2017
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.74	152	2541	0.40	ng/ul	0.00
2) Naphthalene-d8	10.52	136	12708	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.37	164	7678	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.10	188	19236	0.40	ng/ul	0.00
16) Chrysene-d12	21.26	240	18407	0.40	ng/ul	0.00
20) Perylene-d12	23.67	264	16905	0.40	ng/ul	0.00

System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.11	152	792	0.04	ng/ul	0.00
14) Fluoranthene-d10	19.13	212	2290	0.04	ng/ul	0.00

Target Compounds				Qvalue		
3) Naphthalene	10.57	128	1393	0.04	ng/ul#	82
7) Acenaphthylene	14.09	152	4813	0.15	ng/ul	97
12) Phenanthrene	17.13	178	43480	0.82	ng/ul#	86
13) Anthracene	17.24	178	9476	0.21	ng/ul#	86
15) Fluoranthene	19.16	202	78447	1.21	ng/ul	84
17) Pyrene	19.52	202	56728	1.05	ng/ul#	87
18) Benzo(a)anthracene	21.25	228	15295	0.31	ng/ul#	51
19) Chrysene	21.30	228	33777	0.66	ng/ul#	74
21) Benzo(b)fluoranthene	22.94	252	36600m	0.71	ng/ul	
22) Benzo(k)fluoranthene	22.98	252	12560m	0.23	ng/ul	
23) Benzo(a)pyrene	23.56	252	11263	0.23	ng/ul#	46
24) Indeno(1,2,3-cd)pyrene	26.19	276	11717	0.21	ng/ul#	90
26) Benzo(g,h,i)perylene	26.97	276	9759	0.20	ng/ul#	45

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BE093238.D
Acq On : 16 Jun 2017 14:25
Operator : SJ/MA
Sample : I3521-06DL 5X
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
F5C54DL

Manual Integrations
APPROVED

mohammad
6/19/2017 11:07:46 AM

Quant Time: Jun 19 07:30:18 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 16 05:36:31 2017
Response via : Initial Calibration

