

Data Path : Z:\HPCHEM1\BNA_E\Data\BE061517\
 Data File : BE093207.D
 Acq On : 15 Jun 2017 14:21
 Operator : SJ/MA
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.408

Quant Time: Jun 16 04:36:18 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 15 07:58:47 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	59	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	66	0.00
3	Naphthalene	0.992	1.025	-3.3	69	0.00
4 SURR	2-Methylnaphthalene-d10	0.586	0.654	-11.6	74	0.00
5	2-Methylnaphthalene	0.679	0.753	-10.9	74	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	72	0.00
7	Acenaphthylene	1.717	1.975	-15.0	83	0.00
8 C	Acenaphthene	1.385	1.425	-2.9	74	0.00
9	Fluorene	1.624	1.607	1.0	72	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	61	0.00
11	Pentachlorophenol	0.059	0.075	-27.1#	84	0.02
12	Phenanthrene	1.101	1.117	-1.5	64	0.00
13	Anthracene	0.958	1.104	-15.2	72	0.00
14 SURR	Fluoranthene-d10	1.088	1.108	-1.8	63	0.00
15 C	Fluoranthene	1.350	1.346	0.3	62	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	51	0.00
17	Pyrene	1.180	1.409	-19.4	58	0.00
18	Benzo(a)anthracene	1.067	1.226	-14.9	59	0.00
19	Chrysene	1.107	1.132	-2.3	53	0.00
20 I	Perylene-d12	1.000	1.000	0.0	52	0.00
21	Benzo(b)fluoranthene	1.214	1.286	-5.9	57	0.00
22	Benzo(k)fluoranthene	1.289	1.225	5.0	50	0.00
23 C	Benzo(a)pyrene	1.148	1.225	-6.7	58	0.00
24	Indeno(1,2,3-cd)pyrene	1.329	1.389	-4.5	56	0.00
25	Dibenzo(a,h)anthracene	1.119	1.154	-3.1	55	0.00
26	Benzo(g,h,i)perylene	1.140	1.155	-1.3	55	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0