

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE061517\
 Data File : BE093236.D
 Acq On : 16 Jun 2017 10:39
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.411

Quant Time: Jun 16 14:39:51 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 16 05:36:31 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	57	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	62	0.00
3	Naphthalene	0.992	1.022	-3.0	64	0.00
4 SURR	2-Methylnaphthalene-d10	0.586	0.632	-7.8	66	0.00
5	2-Methylnaphthalene	0.679	0.734	-8.1	67	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	63	0.00
7	Acenaphthylene	1.717	1.926	-12.2	71	0.00
8 C	Acenaphthene	1.385	1.433	-3.5	65	0.00
9	Fluorene	1.624	1.607	1.0	63	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	56	0.00
11	Pentachlorophenol	0.059	0.065	-10.2	66	0.02
12	Phenanthrene	1.101	1.111	-0.9	58	0.00
13	Anthracene	0.958	1.044	-9.0	62	0.00
14 SURR	Fluoranthene-d10	1.088	1.104	-1.5	57	0.00
15 C	Fluoranthene	1.350	1.353	-0.2	57	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	48	0.00
17	Pyrene	1.180	1.375	-16.5	53	0.00
18	Benzo(a)anthracene	1.067	1.230	-15.3	56	0.00
19	Chrysene	1.107	1.159	-4.7	51	0.00
20 I	Perylene-d12	1.000	1.000	0.0	49	0.00
21	Benzo(b)fluoranthene	1.214	1.313	-8.2	55	0.00
22	Benzo(k)fluoranthene	1.289	1.248	3.2	49	0.00
23 C	Benzo(a)pyrene	1.148	1.227	-6.9	55	0.00
24	Indeno(1,2,3-cd)pyrene	1.329	1.392	-4.7	53	0.00
25	Dibenzo(a,h)anthracene	1.119	1.157	-3.4	52	0.00
26	Benzo(g,h,i)perylene	1.140	1.149	-0.8	52	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0