

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE061517\
 Data File : BE093241.D
 Acq On : 16 Jun 2017 16:16
 Operator : SJ/MA
 Sample : SSTDC00.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.412

Quant Time: Jun 16 17:01:19 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 16 16:46:59 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	66	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	71	0.00
3	Naphthalene	0.992	1.030	-3.8	75	0.00
4 SURR	2-Methylnaphthalene-d10	0.586	0.637	-8.7	77	0.00
5	2-Methylnaphthalene	0.679	0.733	-8.0	77	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	74	0.00
7	Acenaphthylene	1.717	1.963	-14.3	85	0.00
8 C	Acenaphthene	1.385	1.433	-3.5	76	0.00
9	Fluorene	1.624	1.612	0.7	74	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	63	0.00
11	Pentachlorophenol	0.059	0.074	-25.4#	85	0.02
12	Phenanthrene	1.101	1.119	-1.6	66	0.00
13	Anthracene	0.958	1.122	-17.1	75	0.00
14 SURR	Fluoranthene-d10	1.088	1.147	-5.4	67	0.00
15 C	Fluoranthene	1.350	1.380	-2.2	65	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	54	0.00
17	Pyrene	1.180	1.421	-20.4#	61	0.00
18	Benzo(a)anthracene	1.067	1.233	-15.6	62	0.00
19	Chrysene	1.107	1.142	-3.2	56	0.00
20 I	Perylene-d12	1.000	1.000	0.0	55	0.00
21	Benzo(b)fluoranthene	1.214	1.264	-4.1	60	0.00
22	Benzo(k)fluoranthene	1.289	1.261	2.2	55	0.00
23 C	Benzo(a)pyrene	1.148	1.211	-5.5	61	0.00
24	Indeno(1,2,3-cd)pyrene	1.329	1.380	-3.8	60	0.00
25	Dibenzo(a,h)anthracene	1.119	1.140	-1.9	58	0.00
26	Benzo(g,h,i)perylene	1.140	1.140	0.0	58	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0