

Data Path : Z:\HPCHEM1\BNA_E\Data\BE061616\
 Data File : BE091935.D
 Acq On : 16 Jun 2016 11:19
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 16 15:47:03 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061416.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 15 12:00:38 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	82	0.00
2	1,4-Dioxane	0.705	0.686	2.7	78	0.00
3	n-Nitrosodimethylamine	0.802	0.688	14.2	70	0.03
4 S	2-Fluorophenol	0.977	0.925	5.3	79	0.02
5 S	Phenol-d6	1.267	1.176	7.2	80	0.02
6	bis(2-Chloroethyl)ether	1.431	1.258	12.1	73	0.02
7	n-Nitroso-di-n-propylamine	1.137	1.051	7.6	77	0.02
8 I	Naphthalene-d8	1.000	1.000	0.0	85	0.03
9 S	Nitrobenzene-d5	0.345	0.321	7.0	78	0.02
10	Nitrobenzene	0.343	0.305	11.1	74	0.02
11	bis(2-Chloroethoxy)methane	0.438	0.440	-0.5	79	0.03
12	Naphthalene	1.092	1.062	2.7	77	0.00
13	Hexachlorobutadiene	0.177	0.170	4.0	77	0.00
14	2-Methylnaphthalene	0.741	0.699	5.7	77	0.01
15 I	Acenaphthene-d10	1.000	1.000	0.0	82	0.00
16 S	2,4,6-Tribromophenol	0.102	0.086	15.7	74	0.02
17 S	2-Fluorobiphenyl	1.269	1.280	-0.9	77	0.00
18	Acenaphthylene	5.075	4.640	8.6	76	0.03
19	2,6-Dinitrotoluene	0.811	0.661	18.5	95	0.00
20	Acenaphthene	1.398	1.341	4.1	79	0.00
21	2,4-Dinitrotoluene	0.758	0.627	17.3	73	0.03
22	Fluorene	1.391	1.272	8.6	75	0.02
23	4-Chlorophenyl-phenylether	0.685	0.642	6.3	73	0.02
24 I	Phenanthrene-d10	1.000	1.000	0.0	80	0.00
25	4-Bromophenyl-phenylether	0.193	0.178	7.8	72	0.02
26	Hexachlorobenzene	0.198	0.188	5.1	73	0.00
27	Pentachlorophenol	0.054	0.036	33.3#	67	0.02
28	Phenanthrene	1.250	1.197	4.2	73	0.00
29	Anthracene	1.106	1.007	9.0	74	0.00
30	Fluoranthene	5.354	5.039	5.9	74	0.00
31 I	Chrysene-d12	1.000	1.000	0.0	82	0.00
32	Benzidine	0.170	0.135	20.6	91	0.00
33	Pyrene	3.401	2.967	12.8	73	0.00
34 S	Terphenyl-d14	0.612	0.550	10.1	71	0.00
35	Benzo(a)anthracene	0.786	0.772	1.8	82	0.00
36	3,3'-Dichlorobenzidine	0.127	0.102	19.7	79	0.03
37	Chrysene	1.147	1.141	0.5	73	0.00
38	Bis(2-ethylhexyl)phthalate	0.580	0.562	3.1	92	-0.03
39	Indeno(1,2,3-cd)pyrene	3.925	3.662	6.7	78	0.00
40 I	Perylene-d12	1.000	1.000	0.0	85	-0.01
41	Benzo(b)fluoranthene	1.284	1.118	12.9	71	-0.01

Data Path : Z:\HPCHEM1\BNA_E\Data\BE061616\
Data File : BE091935.D
Acq On : 16 Jun 2016 11:19
Operator : UM/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 16 15:47:03 2016
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061416.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 15 12:00:38 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
42	Benzo(k)fluoranthene	1.175	1.187	-1.0	85	0.00
43 C	Benzo(a)pyrene	1.122	1.060	5.5	80	0.00
44	Dibenzo(a,h)anthracene	1.092	0.997	8.7	77	-0.02
45	Benzo(g,h,i)perylene	4.507	4.176	7.3	78	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0