


Data Path : Z:\HPCHEM1\BNA_E\Data\BE061616\
 Data File : BE091934.D
 Acq On : 16 Jun 2016 10:40
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061416.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Jun 15 12:00:38 2016



AutoFind: Scans 2543, 2544, 2545; Background Corrected with Scan 2535

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	56.2	46730	PASS
68	69	0.00	2	1.6	686	PASS
69	198	0.00	100	50.8	42250	PASS
70	69	0.00	2	0.5	215	PASS
127	198	10	80	53.7	44687	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	83198	PASS
199	198	5	9	7.0	5818	PASS
275	198	10	60	21.7	18042	PASS
365	198	1	100	2.9	2385	PASS
441	442	0.01	24	17.6	8790	PASS
442	198	50	100	60.1	49970	PASS
443	442	15	24	19.1	9562	PASS