

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE061917\
 Data File : BE093259.D
 Acq On : 19 Jun 2017 20:24
 Operator : SJ/MA
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.414

Quant Time: Jun 20 04:20:58 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 20 04:17:17 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	59	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	64	0.00
3	Naphthalene	0.992	1.024	-3.2	67	0.00
4 SURR	2-Methylnaphthalene-d10	0.586	0.646	-10.2	70	0.00
5	2-Methylnaphthalene	0.679	0.746	-9.9	70	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	69	0.00
7	Acenaphthylene	1.717	1.842	-7.3	74	0.00
8 C	Acenaphthene	1.385	1.430	-3.2	71	0.00
9	Fluorene	1.624	1.628	-0.2	70	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	60	0.00
11	Pentachlorophenol	0.059	0.064	-8.5	70	0.00
12	Phenanthrene	1.101	1.125	-2.2	64	0.00
13	Anthracene	0.958	1.037	-8.2	67	0.00
14 SURR	Fluoranthene-d10	1.088	1.086	0.2	61	0.00
15 C	Fluoranthene	1.350	1.309	3.0	59	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	52	0.00
17	Pyrene	1.180	1.343	-13.8	56	0.00
18	Benzo(a)anthracene	1.067	1.179	-10.5	58	0.00
19	Chrysene	1.107	1.151	-4.0	55	0.00
20 I	Perylene-d12	1.000	1.000	0.0	53	0.00
21	Benzo(b)fluoranthene	1.214	1.262	-4.0	57	0.00
22	Benzo(k)fluoranthene	1.289	1.304	-1.2	55	0.00
23 C	Benzo(a)pyrene	1.148	1.219	-6.2	59	0.00
24	Indeno(1,2,3-cd)pyrene	1.329	1.386	-4.3	57	0.00
25	Dibenzo(a,h)anthracene	1.119	1.165	-4.1	57	0.00
26	Benzo(g,h,i)perylene	1.140	1.143	-0.3	56	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0