

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE061917\
 Data File : BE093274.D
 Acq On : 20 Jun 2017 5:39
 Operator : SJ/MA
 Sample : SSTDC00.4EC
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.415

Quant Time: Jun 20 08:51:58 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 20 08:49:11 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	49	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	56	0.00
3	Naphthalene	0.992	1.015	-2.3	58	0.00
4 SURR	2-Methylnaphthalene-d10	0.586	0.633	-8.0	60	0.00
5	2-Methylnaphthalene	0.679	0.731	-7.7	60	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	58	0.00
7	Acenaphthylene	1.717	2.115	-23.2#	72	0.00
8 C	Acenaphthene	1.385	1.430	-3.2	60	0.00
9	Fluorene	1.624	1.669	-2.8	61	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	52	0.00
11	Pentachlorophenol	0.059	0.086	-45.8#	81	0.00
12	Phenanthrene	1.101	1.130	-2.6	55	0.00
13	Anthracene	0.958	1.159	-21.0#	64	0.00
14 SURR	Fluoranthene-d10	1.088	1.155	-6.2	56	0.00
15 C	Fluoranthene	1.350	1.367	-1.3	53	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	44	0.00
17	Pyrene	1.180	1.396	-18.3	50	0.00
18	Benzo(a)anthracene	1.067	1.234	-15.7	51	0.00
19	Chrysene	1.107	1.130	-2.1	46	0.00
20 I	Perylene-d12	1.000	1.000	0.0	46	0.00
21	Benzo(b)fluoranthene	1.214	1.287	-6.0	50	0.00
22	Benzo(k)fluoranthene	1.289	1.220	5.4	44	0.00
23 C	Benzo(a)pyrene	1.148	1.226	-6.8	51	0.00
24	Indeno(1,2,3-cd)pyrene	1.329	1.370	-3.1	49	0.00
25	Dibenzo(a,h)anthracene	1.119	1.142	-2.1	48	0.00
26	Benzo(g,h,i)perylene	1.140	1.131	0.8	48	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0