

Data Path : Z:\HPCHEM1\BNA_E\Data\BE062017\
 Data File : BE093276.D
 Acq On : 20 Jun 2017 11:31
 Operator : SJ/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.416

Quant Time: Jun 21 01:48:19 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 20 08:49:11 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	50	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	54	0.00
3	Naphthalene	0.992	1.020	-2.8	57	0.00
4 SURR	2-Methylnaphthalene-d10	0.586	0.635	-8.4	59	0.00
5	2-Methylnaphthalene	0.679	0.730	-7.5	59	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	57	0.00
7	Acenaphthylene	1.717	1.923	-12.0	64	0.00
8 C	Acenaphthene	1.385	1.421	-2.6	58	0.00
9	Fluorene	1.624	1.623	0.1	58	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	51	0.00
11	Pentachlorophenol	0.059	0.069	-16.9	63	0.00
12	Phenanthrene	1.101	1.124	-2.1	53	0.00
13	Anthracene	0.958	1.061	-10.8	57	0.00
14 SURR	Fluoranthene-d10	1.088	1.126	-3.5	53	0.00
15 C	Fluoranthene	1.350	1.368	-1.3	52	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	45	0.00
17	Pyrene	1.180	1.385	-17.4	50	0.00
18	Benzo(a)anthracene	1.067	1.225	-14.8	51	0.00
19	Chrysene	1.107	1.137	-2.7	46	0.00
20 I	Perylene-d12	1.000	1.000	0.0	46	0.00
21	Benzo(b)fluoranthene	1.214	1.297	-6.8	50	0.00
22	Benzo(k)fluoranthene	1.289	1.252	2.9	45	0.00
23 C	Benzo(a)pyrene	1.148	1.220	-6.3	51	0.00
24	Indeno(1,2,3-cd)pyrene	1.329	1.399	-5.3	50	0.00
25	Dibenzo(a,h)anthracene	1.119	1.168	-4.4	49	0.00
26	Benzo(g,h,i)perylene	1.140	1.151	-1.0	48	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0