

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE062017\
 Data File : BE093306.D
 Acq On : 21 Jun 2017 6:12
 Operator : SJ/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.418

Quant Time: Jun 21 07:03:17 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 21 03:23:49 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	42	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	48	0.00
3	Naphthalene	0.992	1.022	-3.0	50	0.00
4 SURR	2-Methylnaphthalene-d10	0.586	0.642	-9.6	53	0.00
5	2-Methylnaphthalene	0.679	0.742	-9.3	53	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	52	0.00
7	Acenaphthylene	1.717	1.909	-11.2	58	0.00
8 C	Acenaphthene	1.385	1.432	-3.4	54	0.00
9	Fluorene	1.624	1.641	-1.0	53	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	47	0.00
11	Pentachlorophenol	0.059	0.065	-10.2	55	0.00
12	Phenanthrene	1.101	1.128	-2.5	49	0.00
13	Anthracene	0.958	1.088	-13.6	54	0.00
14 SURR	Fluoranthene-d10	1.088	1.092	-0.4	48	0.00
15 C	Fluoranthene	1.350	1.321	2.1	46	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	40	0.00
17	Pyrene	1.180	1.393	-18.1	44	0.00
18	Benzo(a)anthracene	1.067	1.221	-14.4	45	0.00
19	Chrysene	1.107	1.135	-2.5	41	0.00
20 I	Perylene-d12	1.000	1.000	0.0	40	0.00
21	Benzo(b)fluoranthene	1.214	1.280	-5.4	44	0.00
22	Benzo(k)fluoranthene	1.289	1.279	0.8	41	0.00
23 C	Benzo(a)pyrene	1.148	1.217	-6.0	44	0.00
24	Indeno(1,2,3-cd)pyrene	1.329	1.391	-4.7	44	0.00
25	Dibenzo(a,h)anthracene	1.119	1.161	-3.8	43	0.00
26	Benzo(g,h,i)perylene	1.140	1.146	-0.5	42	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0