

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE062017\
 Data File : BE093290.D
 Acq On : 20 Jun 2017 20:25
 Operator : SJ/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTD0.417

Quant Time: Jun 21 01:51:43 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 21 01:49:50 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.74	152	1808	0.40	ng/ul	0.00
2) Naphthalene-d8	10.52	136	9344	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.36	164	5412	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.10	188	13883	0.40	ng/ul	0.00
16) Chrysene-d12	21.26	240	14131	0.40	ng/ul	0.00
20) Perylene-d12	23.66	264	12078	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.11	152	6012	0.44	ng/ul	0.00
14) Fluoranthene-d10	19.13	212	14999	0.40	ng/ul	0.00
Target Compounds					Qvalue	
3) Naphthalene	10.56	128	9577	0.413	ng/ul#	92
5) 2-Methylnaphthalene	12.18	142	6850	0.432	ng/ul	99
7) Acenaphthylene	14.09	152	10066	0.433	ng/ul	96
8) Acenaphthene	14.42	153	7735	0.413	ng/ul	98
9) Fluorene	15.42	166	8839	0.402	ng/ul	95
11) Pentachlorophenol	16.75	266	901	0.437	ng/ul	96
12) Phenanthrene	17.13	178	15583	0.408	ng/ul#	89
13) Anthracene	17.22	178	14568	0.438	ng/ul#	89
15) Fluoranthene	19.15	202	18401	0.393	ng/ul	84
17) Pyrene	19.51	202	19848	0.476	ng/ul#	86
18) Benzo(a)anthracene	21.24	228	16979	0.450	ng/ul#	85
19) Chrysene	21.30	228	16342	0.418	ng/ul	98
21) Benzo(b)fluoranthene	22.93	252	15136	0.413	ng/ul	88
22) Benzo(k)fluoranthene	22.97	252	15422	0.396	ng/ul#	89
23) Benzo(a)pyrene	23.55	252	14595	0.421	ng/ul#	83
24) Indeno(1,2,3-cd)pyrene	26.18	276	16829	0.419	ng/ul#	94
25) Dibenzo(a,h)anthracene	26.20	278	13951	0.413	ng/ul#	89
26) Benzo(g,h,i)perylene	26.96	276	13603	0.395	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE062017\
Data File : BE093290.D
Acq On : 20 Jun 2017 20:25
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTD0.417

Quant Time: Jun 21 01:51:43 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 21 01:49:50 2017
Response via : Initial Calibration

