

Data Path : Z:\HPCHEM1\BNA E\Data\BE062117\
 Data File : BE093308.D
 Acq On : 21 Jun 2017 9:44
 Operator : SJ/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.419

Quant Time: Jun 22 02:08:33 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE060717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 21 03:23:49 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	47	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	53	0.00
3	Naphthalene	0.992	1.024	-3.2	56	0.00
4 SURR	2-Methylnaphthalene-d10	0.586	0.637	-8.7	57	0.00
5	2-Methylnaphthalene	0.679	0.736	-8.4	58	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	55	0.00
7	Acenaphthylene	1.717	1.861	-8.4	60	0.00
8 C	Acenaphthene	1.385	1.431	-3.3	57	0.00
9	Fluorene	1.624	1.614	0.6	56	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	47	0.00
11	Pentachlorophenol	0.059	0.059	0.0	51	0.00
12	Phenanthrene	1.101	1.125	-2.2	50	0.00
13	Anthracene	0.958	1.059	-10.5	53	0.00
14 SURR	Fluoranthene-d10	1.088	1.077	1.0	47	0.00
15 C	Fluoranthene	1.350	1.321	2.1	47	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	40	0.00
17	Pyrene	1.180	1.403	-18.9	45	0.00
18	Benzo(a)anthracene	1.067	1.207	-13.1	45	0.00
19	Chrysene	1.107	1.155	-4.3	42	0.00
20 I	Perylene-d12	1.000	1.000	0.0	40	0.00
21	Benzo(b)fluoranthene	1.214	1.287	-6.0	44	0.00
22	Benzo(k)fluoranthene	1.289	1.280	0.7	40	0.00
23 C	Benzo(a)pyrene	1.148	1.223	-6.5	44	0.00
24	Indeno(1,2,3-cd)pyrene	1.329	1.393	-4.8	43	0.00
25	Dibenzo(a,h)anthracene	1.119	1.158	-3.5	42	0.00
26	Benzo(a,h,i)perylene	1.140	1.144	-0.4	42	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0