

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE062118\
 Data File : BE096836.D
 Acq On : 21 Jun 2018 9:26
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jun 22 02:28:20 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE062018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 20 15:45:14 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.599	0.626	-4.5	101	0.00
3	n-Nitrosodimethylamine	0.635	0.661	-4.1	109	0.00
4 S	2-Fluorophenol	0.894	0.874	2.2	106	0.00
5 S	Phenol-d6	1.334	1.263	5.3	104	0.00
6	bis(2-Chloroethyl)ether	1.462	1.416	3.1	101	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	104	0.00
8 S	Nitrobenzene-d5	0.242	0.223	7.9	103	0.00
9	Naphthalene	3.630	3.570	1.7	103	0.00
10	Hexachlorobutadiene	0.151	0.146	3.3	102	0.00
11 SURR	2-Methylnaphthalene-d10	0.576	0.558	3.1	101	0.00
12	2-Methylnaphthalene	0.602	0.581	3.5	102	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
14 S	2,4,6-Tribromophenol	0.095	0.080	15.8	99	0.01
15 S	2-Fluorobiphenyl	1.464	1.507	-2.9	102	0.00
16	Acenaphthylene	1.734	1.721	0.7	102	0.00
17	Acenaphthene	1.147	1.153	-0.5	100	0.00
18	Fluorene	1.348	1.359	-0.8	104	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	104	0.00
20	4-Bromophenyl-phenylether	0.180	0.177	1.7	103	0.00
21	Hexachlorobenzene	0.206	0.203	1.5	100	0.00
22	Pentachlorophenol	0.038	0.031	18.4	109	0.00
23	Phenanthrene	0.983	0.971	1.2	103	0.00
24	Anthracene	0.815	0.782	4.0	105	0.00
25 SURR	Fluoranthene-d10	3.521	3.410	3.2	106	0.00
26	Fluoranthene	0.994	1.013	-1.9	105	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	107	0.00
28	Pyrene	1.157	1.118	3.4	105	0.00
29 S	Terphenyl-d14	0.772	0.752	2.6	106	0.00
30	Benzo(a)anthracene	0.886	0.820	7.4	108	0.00
31	Chrysene	1.095	1.055	3.7	105	0.00
32	Bis(2-ethylhexyl)phthalate	0.794	0.674	15.1	102	0.00
33	Indeno(1,2,3-cd)pyrene	0.814	0.749	8.0	107	0.00
34 I	Perylene-d12	1.000	1.000	0.0	109	0.00
35	Benzo(b)fluoranthene	1.092	1.017	6.9	107	0.00
36	Benzo(k)fluoranthene	1.233	1.165	5.5	110	0.00
37 C	Benzo(a)pyrene	0.867	0.835	3.7	111	0.00
38	Dibenzo(a,h)anthracene	0.889	0.778	12.5	105	0.00
39	Benzo(g,h,i)perylene	0.969	0.876	9.6	106	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE062118\
Data File : BE096836.D
Acq On : 21 Jun 2018 9:26
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Jun 22 02:28:20 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE062018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 20 15:45:14 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		