

Data Path : Z:\HPCHEM1\BNA_E\Data\BE062217\
 Data File : BE093325.D
 Acq On : 22 Jun 2017 12:23
 Operator : SJ/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Jun 23 07:50:06 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 12 14:07:54 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	31#	-0.02
2	1,4-Dioxane	0.887	0.813	8.3	34#	0.00
3	n-Nitrosodimethylamine	0.758	0.800	-5.5	36#	0.00
4 S	2-Fluorophenol	1.180	1.480	-25.4#	45#	0.00
5 S	Phenol-d6	1.729	2.218	-28.3#	45#	0.00
6	bis(2-Chloroethyl)ether	1.510	1.684	-11.5	39#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	36#	-0.02
8 S	Nitrobenzene-d5	0.319	0.387	-21.3	50	0.00
9	Naphthalene	1.073	1.153	-7.5	44#	0.00
10	Hexachlorobutadiene	0.146	0.150	-2.7	43#	-0.02
11 SURR	2-Methylnaphthalene-d10	0.601	0.700	-16.5	49#	-0.01
12	2-Methylnaphthalene	0.666	0.750	-12.6	46#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	41#	0.00
14 S	2,4,6-Tribromophenol	0.091	0.115	-26.4#	63	0.00
15 S	2-Fluorobiphenyl	1.708	1.747	-2.3	47#	0.00
16	Acenaphthylene	6.960	7.770	-11.6	55	0.00
17	Acenaphthene	1.297	1.356	-4.5	50#	-0.02
18	Fluorene	1.563	1.647	-5.4	50	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	33#	-0.03
20	4-Bromophenyl-phenylether	0.244	0.192	21.3	30#	0.00
21	Hexachlorobenzene	0.214	0.272	-27.1#	54	0.00
22	Pentachlorophenol	0.069	0.082	-18.8	49#	0.00
23	Phenanthrene	1.808	1.735	4.0	37#	0.00
24	Anthracene	1.532	1.731	-13.0	48#	0.00
25 SURR	Fluoranthene-d10	1.318	1.365	-3.6	40#	0.00
26	Fluoranthene	6.567	6.679	-1.7	41#	-0.02
27 I	Chrysene-d12	1.000	1.000	0.0	28#	-0.02
28	Pyrene	4.870	5.533	-13.6	35#	0.00
29 S	Terphenyl-d14	0.772	0.900	-16.6	36#	0.00
30	Benzo(a)anthracene	1.080	1.210	-12.0	36#	0.00
31	Chrysene	1.153	1.208	-4.8	33#	0.00
32	Bis(2-ethylhexyl)phthalate	0.525	0.543	-3.4	37#	-0.02
33	Indeno(1,2,3-cd)pyrene	4.306	4.580	-6.4	34#	0.00
34 I	Perylene-d12	1.000	1.000	0.0	27#	0.00
35	Benzo(b)fluoranthene	1.285	1.502	-16.9	37#	0.00
36	Benzo(k)fluoranthene	1.240	1.260	-1.6	31#	0.00
37 C	Benzo(a)pyrene	1.111	1.243	-11.9	35#	0.00
38	Dibenzo(a,h)anthracene	1.158	1.232	-6.4	33#	0.00
39	Benzo(g,h,i)perylene	4.678	4.924	-5.3	33#	0.02

Data Path : Z:\HPCHEM1\BNA_E\Data\BE062217\
Data File : BE093325.D
Acq On : 22 Jun 2017 12:23
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 23 07:50:06 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 12 14:07:54 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			