

Data Path : Z:\HPCHEM1\BNA_E\Data\BE062217\
 Data File : BE093332.D
 Acq On : 22 Jun 2017 17:25
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 23 08:10:53 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 12 14:07:54 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	28#	-0.02
2	1,4-Dioxane	0.887	0.782	11.8	30#	-0.01
3	n-Nitrosodimethylamine	0.758	0.738	2.6	30#	-0.01
4 S	2-Fluorophenol	1.180	1.330	-12.7	37#	0.00
5 S	Phenol-d6	1.729	2.175	-25.8#	41#	0.00
6	bis(2-Chloroethyl)ether	1.510	1.549	-2.6	33#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	32#	-0.02
8 S	Nitrobenzene-d5	0.319	0.367	-15.0	43#	-0.02
9	Naphthalene	1.073	1.126	-4.9	38#	0.00
10	Hexachlorobutadiene	0.146	0.150	-2.7	38#	-0.02
11 SURR	2-Methylnaphthalene-d10	0.601	0.700	-16.5	43#	-0.01
12	2-Methylnaphthalene	0.666	0.761	-14.3	42#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	39#	0.00
14 S	2,4,6-Tribromophenol	0.091	0.122	-34.1#	62	0.00
15 S	2-Fluorobiphenyl	1.708	1.718	-0.6	43#	0.00
16	Acenaphthylene	6.960	7.867	-13.0	52	0.00
17	Acenaphthene	1.297	1.373	-5.9	47#	-0.02
18	Fluorene	1.563	1.683	-7.7	48#	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	35#	-0.03
20	4-Bromophenyl-phenylether	0.244	0.182	25.4#	30#	0.00
21	Hexachlorobenzene	0.214	0.277	-29.4#	58	0.00
22	Pentachlorophenol	0.069	0.082	-18.8	52	0.00
23	Phenanthrene	1.808	1.469	18.7	33#	0.00
24	Anthracene	1.532	1.525	0.5	45#	0.00
25 SURR	Fluoranthene-d10	1.318	1.361	-3.3	43#	0.00
26	Fluoranthene	6.567	6.396	2.6	42#	-0.02
27 I	Chrysene-d12	1.000	1.000	0.0	28#	-0.02
28	Pyrene	4.870	5.270	-8.2	33#	0.00
29 S	Terphenyl-d14	0.772	0.844	-9.3	34#	0.00
30	Benzo(a)anthracene	1.080	1.276	-18.1	37#	0.00
31	Chrysene	1.153	1.237	-7.3	33#	0.00
32	Bis(2-ethylhexyl)phthalate	0.525	0.547	-4.2	37#	-0.02
33	Indeno(1,2,3-cd)pyrene	4.306	4.579	-6.3	34#	0.00
34 I	Perylene-d12	1.000	1.000	0.0	26#	-0.01
35	Benzo(b)fluoranthene	1.285	1.356	-5.5	32#	-0.01
36	Benzo(k)fluoranthene	1.240	1.405	-13.3	34#	-0.01
37 C	Benzo(a)pyrene	1.111	1.232	-10.9	34#	0.00
38	Dibenzo(a,h)anthracene	1.158	1.242	-7.3	33#	0.00
39	Benzo(g,h,i)perylene	4.678	4.942	-5.6	32#	0.02

Data Path : Z:\HPCHEM1\BNA_E\Data\BE062217\
Data File : BE093332.D
Acq On : 22 Jun 2017 17:25
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 23 08:10:53 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 12 14:07:54 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			