


Data Path : Z:\HPCHEM1\BNA_E\Data\BE062217\
Data File : BE093324.D
Acq On : 22 Jun 2017 11:00
Operator : SJ/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE061217.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Mon Jun 12 14:07:54 2017



AutoFind: Scans 2547, 2548, 2549; Background Corrected with Scan 2540

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	50.5	10603	PASS
68	69	0.00	2	0.4	40	PASS
69	198	0.00	100	53.2	11169	PASS
70	69	0.00	2	0.5	57	PASS
127	198	10	80	57.4	12042	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	20986	PASS
199	198	5	9	6.4	1336	PASS
275	198	10	60	24.8	5206	PASS
365	198	1	100	2.9	616	PASS
441	442	0.01	24	15.6	2605	PASS
442	198	50	100	79.5	16678	PASS
443	442	15	24	20.1	3350	PASS