

Data Path : Z:\HPCHEM1\BNA_E\Data\BE062315\
 Data File : BE089931.D
 Acq On : 24 Jun 2015 1:35
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICV001

Quant Time: Jun 24 09:30:59 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 24 09:11:03 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	83	0.00
2	1,4-Dioxane	1.000	1.130	-13.0	92	0.00
3	n-Nitrosodimethylamine	1.000	1.016	-1.6	89	0.00
4 S	2-Fluorophenol	1.000	0.715	28.5#	79	0.00
5 S	Phenol-d6	1.000	0.846	15.4	81	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	81	0.00
7 S	Nitrobenzene-d5	1.000	0.921	7.9	86	0.00
8	Nitrobenzene	1.000	0.915	8.5	87	0.00
9	Naphthalene	1.000	0.913	8.7	86	0.00
10	Hexachlorobutadiene	1.000	0.868	13.2	87	0.00
11	2-Methylnaphthalene	1.000	0.962	3.8	87	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	81	0.00
13 S	2,4,6-Tribromophenol	1.000	0.355	64.5#	76	0.00
14 S	2-Fluorobiphenyl	1.000	0.823	17.7	88	0.00
15	Acenaphthylene	1.000	2.756	-175.6#	85	0.00
16	Acenaphthene	1.000	0.853	14.7	83	0.00
17	Fluorene	1.000	1.302	-30.2#	85	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	80	0.00
19	4-Bromophenyl-phenylether	1.000	0.978	2.2	90	0.00
20	Hexachlorobenzene	1.000	3.787	-278.7#	89	0.00
21	Pentachlorophenol	1.000	0.456	54.4#	83	0.00
22	Phenanthrene	1.000	0.744	25.6#	87	0.00
23	Anthracene	1.000	1.502	-50.2#	87	0.00
24	Fluoranthene	1.000	1.221	-22.1	88	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	83	0.00
26	Pyrene	1.000	0.742	25.8#	87	0.00
27 S	Terphenyl-d14	1.000	1.152	-15.2	91	0.00
28	Benzo(a)anthracene	1.000	0.783	21.7	87	0.00
29	Chrysene	1.000	0.830	17.0	90	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	0.121	87.9#	90	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	0.630	37.0#	84	0.00
32 I	Perylene-d12	5.000	5.000	0.0	82	0.00
33	Benzo(b)fluoranthene	1.000	0.659	34.1#	85	0.00
34	Benzo(k)fluoranthene	1.000	0.839	16.1	88	0.00
35 C	Benzo(a)pyrene	1.000	0.685	31.5#	85	0.00
36	Dibenzo(a,h)anthracene	1.000	0.650	35.0#	86	0.00
37	Benzo(g,h,i)perylene	1.000	0.658	34.2#	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1