

Data Path : Z:\HPCHEM1\BNA_E\Data\BE062315\
 Data File : BE089931.D
 Acq On : 24 Jun 2015 1:35
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICV001

Quant Time: Jun 24 09:30:59 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 24 09:11:03 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	0.00
2	1,4-Dioxane	0.675	0.641	5.0	92	0.00
3	n-Nitrosodimethylamine	0.899	0.884	1.7	89	0.00
4 S	2-Fluorophenol	0.903	0.807	10.6	79	0.00
5 S	Phenol-d6	1.383	1.251	9.5	81	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	81	0.00
7 S	Nitrobenzene-d5	0.372	0.354	4.8	86	0.00
8	Nitrobenzene	0.396	0.377	4.8	87	0.00
9	Naphthalene	1.112	1.095	1.5	86	0.00
10	Hexachlorobutadiene	0.176	0.174	1.1	87	0.00
11	2-Methylnaphthalene	0.761	0.739	2.9	87	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	81	0.00
13 S	2,4,6-Tribromophenol	0.064	0.050	21.9	76	0.00
14 S	2-Fluorobiphenyl	1.457	1.421	2.5	88	0.00
15	Acenaphthylene	8.767	8.558	2.4	85	0.00
16	Acenaphthene	1.288	1.244	3.4	83	0.00
17	Fluorene	1.750	1.697	3.0	85	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	80	0.00
19	4-Bromophenyl-phenylether	0.204	0.199	2.5	90	0.00
20	Hexachlorobenzene	0.894	0.899	-0.6	89	0.00
21	Pentachlorophenol	0.033	0.026	21.2	83	0.00
22	Phenanthrene	1.254	1.228	2.1	87	0.00
23	Anthracene	1.164	1.149	1.3	87	0.00
24	Fluoranthene	1.361	1.354	0.5	88	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	83	0.00
26	Pyrene	1.266	1.250	1.3	87	0.00
27 S	Terphenyl-d14	0.863	0.860	0.3	91	0.00
28	Benzo(a)anthracene	1.195	1.144	4.3	87	0.00
29	Chrysene	1.306	1.275	2.4	90	0.00
30	Bis(2-ethylhexyl)phthalate	0.102	0.093	8.8	90	0.00
31	Indeno(1,2,3-cd)pyrene	0.960	0.915	4.7	84	0.00
32 I	Perylene-d12	1.000	1.000	0.0	82	0.00
33	Benzo(b)fluoranthene	1.058	1.033	2.4	85	0.00
34	Benzo(k)fluoranthene	1.199	1.205	-0.5	88	0.00
35 C	Benzo(a)pyrene	0.940	0.907	3.5	85	0.00
36	Dibenzo(a,h)anthracene	0.875	0.848	3.1	86	0.00
37	Benzo(g,h,i)perylene	0.941	0.905	3.8	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0