

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA_E\Data\BE062317\
 Data File : BE093340.D
 Acq On : 23 Jun 2017 16:56
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.422

Quant Time: Jun 24 03:59:16 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jun 24 03:58:33 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	47	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	52	0.00
3	Naphthalene	0.992	0.982	1.0	53	0.00
4 SURR	2-Methylnaphthalene-d10	0.586	0.615	-4.9	55	0.00
5	2-Methylnaphthalene	0.679	0.713	-5.0	55	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	56	0.00
7	Acenaphthylene	1.717	1.792	-4.4	59	0.00
8 C	Acenaphthene	1.385	1.387	-0.1	56	0.00
9	Fluorene	1.624	1.579	2.8	55	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	50	0.00
11	Pentachlorophenol	0.059	0.066	-11.9	60	0.00
12	Phenanthrene	1.101	1.109	-0.7	52	0.00
13	Anthracene	0.958	1.069	-11.6	56	0.00
14 SURR	Fluoranthene-d10	1.088	1.083	0.5	50	0.00
15 C	Fluoranthene	1.350	1.299	3.8	48	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	43	0.00
17	Pyrene	1.180	1.323	-12.1	46	0.00
18	Benzo(a)anthracene	1.067	1.164	-9.1	47	0.00
19	Chrysene	1.107	1.082	2.3	43	0.00
20 I	Perylene-d12	1.000	1.000	0.0	43	0.00
21	Benzo(b)fluoranthene	1.214	1.253	-3.2	46	0.00
22	Benzo(k)fluoranthene	1.289	1.246	3.3	43	0.00
23 C	Benzo(a)pyrene	1.148	1.192	-3.8	47	0.00
24	Indeno(1,2,3-cd)pyrene	1.329	1.354	-1.9	46	0.00
25	Dibenzo(a,h)anthracene	1.119	1.124	-0.4	44	0.00
26	Benzo(g,h,i)perylene	1.140	1.090	4.4	43	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0