

Data Path : Z:\HPCHEM1\BNA_E\Data\BE062415\
 Data File : BE089933.D
 Acq On : 24 Jun 2015 14:30
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jun 25 03:45:05 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 24 09:27:42 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.675	0.564	16.4	99	0.00
3	n-Nitrosodimethylamine	0.899	0.801	10.9	98	0.00
4 S	2-Fluorophenol	0.903	0.749	17.1	89	0.00
5 S	Phenol-d6	1.383	1.176	15.0	92	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	101	0.00
7 S	Nitrobenzene-d5	0.372	0.332	10.8	101	0.00
8	Nitrobenzene	0.396	0.351	11.4	101	0.00
9	Naphthalene	1.112	1.030	7.4	101	0.00
10	Hexachlorobutadiene	0.176	0.163	7.4	101	0.00
11	2-Methylnaphthalene	0.761	0.678	10.9	99	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
13 S	2,4,6-Tribromophenol	0.064	0.044	31.3#	82	0.00
14 S	2-Fluorobiphenyl	1.457	1.310	10.1	99	0.00
15	Acenaphthylene	8.767	8.092	7.7	98	0.00
16	Acenaphthene	1.288	1.207	6.3	98	0.00
17	Fluorene	1.750	1.583	9.5	97	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	94	0.00
19	4-Bromophenyl-phenylether	0.204	0.182	10.8	96	0.00
20	Hexachlorobenzene	0.894	0.828	7.4	97	0.00
21	Pentachlorophenol	0.033	0.022	33.3#	83	-0.02
22	Phenanthrene	1.254	1.158	7.7	97	0.00
23	Anthracene	1.164	1.061	8.8	95	0.00
24	Fluoranthene	1.361	1.257	7.6	96	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	98	0.00
26	Pyrene	1.266	1.162	8.2	95	0.00
27 S	Terphenyl-d14	0.863	0.769	10.9	97	0.00
28	Benzo(a)anthracene	1.195	1.080	9.6	97	0.00
29	Chrysene	1.306	1.163	10.9	98	0.00
30	Bis(2-ethylhexyl)phthalate	0.102	0.080	21.6	92	0.00
31	Indeno(1,2,3-cd)pyrene	0.960	0.886	7.7	96	0.00
32 I	Perylene-d12	1.000	1.000	0.0	99	0.00
33	Benzo(b)fluoranthene	1.058	0.998	5.7	98	0.00
34	Benzo(k)fluoranthene	1.199	1.101	8.2	96	0.00
35 C	Benzo(a)pyrene	0.940	0.853	9.3	96	0.00
36	Dibenzo(a,h)anthracene	0.875	0.792	9.5	96	0.00
37	Benzo(g,h,i)perylene	0.941	0.864	8.2	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0